

CALIBRATION VERIFICATION SUMMARY

Contract: BAPS03

Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 11/13/2024

Lab Sample No.: PSTDCCC050 Data File : PL093019.D Time Analyzed: 09:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.712	6.610	6.810	46.590	50.000	-6.8
4,4'-DDE	6.195	6.093	6.293	46.170	50.000	-7.7
4,4'-DDT	7.025	6.923	7.123	45.880	50.000	-8.2
Aldrin	5.260	5.158	5.358	46.340	50.000	-7.3
alpha-BHC	3.998	3.896	4.096	47.960	50.000	-4.1
alpha-Chlordane	6.022	5.919	6.119	45.810	50.000	-8.4
beta-BHC	4.528	4.425	4.625	47.740	50.000	-4.5
Decachlorobiphenyl	9.056	8.954	9.154	44.410	50.000	-11.2
delta-BHC	4.775	4.672	4.872	47.660	50.000	-4.7
Dieldrin	6.347	6.245	6.445	45.310	50.000	-9.4
Endosulfan I	6.072	5.970	6.170	45.430	50.000	-9.1
Endosulfan II	6.796	6.694	6.894	43.220	50.000	-13.6
Endosulfan sulfate	7.161	7.058	7.258	44.520	50.000	-11.0
Endrin	6.576	6.474	6.674	44.400	50.000	-11.2
Endrin aldehyde	6.927	6.824	7.024	45.060	50.000	-9.9
Endrin ketone	7.645	7.543	7.743	44.210	50.000	-11.6
gamma-BHC (Lindane)	4.331	4.228	4.428	47.750	50.000	-4.5
gamma-Chlordane	5.943	5.841	6.041	45.840	50.000	-8.3
Heptachlor	4.919	4.817	5.017	47.040	50.000	-5.9
Heptachlor epoxide	5.685	5.584	5.784	46.570	50.000	-6.9
Methoxychlor	7.501	7.399	7.599	45.940	50.000	-8.1
Tetrachloro-m-xylene	3.542	3.440	3.640	48.480	50.000	-3.0

CALIBRATION VERIFICATION SUMMARY

Contract: BAPS03

Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 11/13/2024

Lab Sample No.: PSTDCCC050 Data File : PL093019.D Time Analyzed: 09:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.790	5.689	5.889	56.030	50.000	12.1
4,4'-DDE	5.235	5.135	5.335	55.160	50.000	10.3
4,4'-DDT	6.040	5.940	6.140	53.440	50.000	6.9
Aldrin	4.230	4.129	4.329	54.550	50.000	9.1
alpha-BHC	3.281	3.181	3.381	53.950	50.000	7.9
alpha-Chlordane	5.046	4.945	5.145	53.820	50.000	7.6
beta-BHC	3.911	3.810	4.010	53.600	50.000	7.2
Decachlorobiphenyl	7.916	7.816	8.016	47.790	50.000	-4.4
delta-BHC	4.140	4.039	4.239	54.990	50.000	10.0
Dieldrin	5.366	5.266	5.466	54.290	50.000	8.6
Endosulfan I	5.102	5.002	5.202	46.890	50.000	-6.2
Endosulfan II	5.936	5.836	6.036	53.540	50.000	7.1
Endosulfan sulfate	6.339	6.238	6.438	52.300	50.000	4.6
Endrin	5.642	5.541	5.741	54.810	50.000	9.6
Endrin aldehyde	6.116	6.015	6.215	51.910	50.000	3.8
Endrin ketone	6.844	6.743	6.943	50.710	50.000	1.4
gamma-BHC (Lindane)	3.611	3.511	3.711	54.010	50.000	8.0
gamma-Chlordane	4.982	4.882	5.082	54.110	50.000	8.2
Heptachlor	3.950	3.849	4.049	54.150	50.000	8.3
Heptachlor epoxide	4.732	4.632	4.832	53.790	50.000	7.6
Methoxychlor	6.615	6.515	6.715	52.250	50.000	4.5
Tetrachloro-m-xylene	2.779	2.678	2.878	52.110	50.000	4.2