

CALIBRATION VERIFICATION SUMMARY

Contract: BAPS03

Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/13/2024

Lab Sample No.: PSTDCCC050 Data File : PL093039.D Time Analyzed: 15:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.708	6.610	6.810	48.710	50.000	-2.6
4,4'-DDE	6.191	6.093	6.293	47.130	50.000	-5.7
4,4'-DDT	7.023	6.923	7.123	43.940	50.000	-12.1
Aldrin	5.257	5.158	5.358	46.740	50.000	-6.5
alpha-BHC	3.994	3.896	4.096	48.840	50.000	-2.3
alpha-Chlordane	6.018	5.919	6.119	46.030	50.000	-7.9
beta-BHC	4.524	4.425	4.625	48.290	50.000	-3.4
Decachlorobiphenyl	9.052	8.954	9.154	44.920	50.000	-10.2
delta-BHC	4.772	4.672	4.872	48.250	50.000	-3.5
Dieldrin	6.343	6.245	6.445	45.840	50.000	-8.3
Endosulfan I	6.068	5.970	6.170	45.860	50.000	-8.3
Endosulfan II	6.793	6.694	6.894	44.940	50.000	-10.1
Endosulfan sulfate	7.157	7.058	7.258	45.200	50.000	-9.6
Endrin	6.574	6.474	6.674	43.880	50.000	-12.2
Endrin aldehyde	6.923	6.824	7.024	45.670	50.000	-8.7
Endrin ketone	7.642	7.543	7.743	45.510	50.000	-9.0
gamma-BHC (Lindane)	4.326	4.228	4.428	48.610	50.000	-2.8
gamma-Chlordane	5.939	5.841	6.041	46.060	50.000	-7.9
Heptachlor	4.915	4.817	5.017	46.490	50.000	-7.0
Heptachlor epoxide	5.682	5.584	5.784	47.320	50.000	-5.4
Methoxychlor	7.498	7.399	7.599	44.300	50.000	-11.4
Tetrachloro-m-xylene	3.538	3.440	3.640	49.310	50.000	-1.4

CALIBRATION VERIFICATION SUMMARY

Contract: BAPS03

Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/13/2024

Lab Sample No.: PSTDCCC050 Data File : PL093039.D Time Analyzed: 15:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.787	5.689	5.889	55.650	50.000	11.3
4,4'-DDE	5.232	5.135	5.335	54.000	50.000	8.0
4,4'-DDT	6.037	5.940	6.140	48.980	50.000	-2.0
Aldrin	4.227	4.129	4.329	53.380	50.000	6.8
alpha-BHC	3.278	3.181	3.381	53.270	50.000	6.5
alpha-Chlordane	5.043	4.945	5.145	52.380	50.000	4.8
beta-BHC	3.908	3.810	4.010	52.980	50.000	6.0
Decachlorobiphenyl	7.913	7.816	8.016	47.140	50.000	-5.7
delta-BHC	4.137	4.039	4.239	55.100	50.000	10.2
Dieldrin	5.364	5.266	5.466	52.500	50.000	5.0
Endosulfan I	5.099	5.002	5.202	46.180	50.000	-7.6
Endosulfan II	5.932	5.836	6.036	52.970	50.000	5.9
Endosulfan sulfate	6.336	6.238	6.438	52.350	50.000	4.7
Endrin	5.638	5.541	5.741	51.820	50.000	3.6
Endrin aldehyde	6.113	6.015	6.215	50.890	50.000	1.8
Endrin ketone	6.840	6.743	6.943	52.320	50.000	4.6
gamma-BHC (Lindane)	3.608	3.511	3.711	52.820	50.000	5.6
gamma-Chlordane	4.980	4.882	5.082	52.630	50.000	5.3
Heptachlor	3.947	3.849	4.049	51.830	50.000	3.7
Heptachlor epoxide	4.728	4.632	4.832	52.680	50.000	5.4
Methoxychlor	6.612	6.515	6.715	49.140	50.000	-1.7
Tetrachloro-m-xylene	2.775	2.678	2.878	51.640	50.000	3.3