

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4822

Client: BAPS North Bergen

Analytical Method: 8081B

DataFile : PL093029.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID: P4821-05MSD	BP-F-2MSD											
	alpha-BHC	18.79	0	20.1	ug/kg	107		2		30 (60)	150 (144)	30 (20)
	beta-BHC	18.79	0	20.9	ug/kg	111		1		30 (54)	150 (143)	30 (20)
	delta-BHC	18.79	0	16.9	ug/kg	90		1		30 (47)	150 (144)	30 (20)
	gamma-BHC (Lindane)	18.79	0	19.8	ug/kg	105		1		30 (61)	150 (140)	30 (20)
	Heptachlor	18.79	0	21.1	ug/kg	112		0		30 (63)	150 (135)	30 (20)
	Aldrin	18.79	0	20.6	ug/kg	110		1		30 (49)	150 (139)	30 (20)
	Heptachlor epoxide	18.79	0	20.8	ug/kg	111		1		30 (32)	150 (180)	30 (20)
	Endosulfan I	18.79	0	21.3	ug/kg	113		0		30 (56)	150 (142)	30 (20)
	Dieldrin	18.79	0	21.3	ug/kg	113		1		30 (47)	150 (161)	30 (20)
	4,4'-DDE	18.79	0	22.1	ug/kg	118		1		30 (55)	150 (136)	30 (20)
	Endrin	18.79	0	22.5	ug/kg	120		1		30 (57)	150 (139)	30 (20)
	Endosulfan II	18.79	0	21.8	ug/kg	116		1		30 (40)	150 (163)	30 (20)
	4,4'-DDD	18.79	0	22.0	ug/kg	117		2		30 (37)	150 (192)	30 (20)
	Endosulfan sulfate	18.79	0	20.7	ug/kg	110		1		30 (62)	150 (139)	30 (20)
	4,4'-DDT	18.79	0	22.1	ug/kg	118		1		30 (51)	150 (146)	30 (20)
	Methoxychlor	18.79	0	21.5	ug/kg	114		0		30 (54)	150 (136)	30 (20)
	Endrin ketone	18.79	0	20.9	ug/kg	111		3		30 (60)	150 (129)	30 (20)
	Endrin aldehyde	18.79	0	20.6	ug/kg	110		2		30 (59)	150 (132)	30 (20)
	alpha-Chlordane	18.79	0	21.8	ug/kg	116		1		30 (30)	150 (192)	30 (20)
	gamma-Chlordane	18.79	0	21.5	ug/kg	114		1		30 (44)	150 (175)	30 (20)