

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112224\
 Data File : BF140564.D
 Acq On : 22 Nov 2024 09:20
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 22 10:52:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	92	0.00
2	1,4-Dioxane	40.000	39.008	2.5	90	0.00
3	Pyridine	40.000	41.129	-2.8	86	0.00
4	n-Nitrosodimethylamine	40.000	38.644	3.4	87	0.00
5 S	2-Fluorophenol	80.000	77.548	3.1	89	0.00
6	Aniline	40.000	40.492	-1.2	81	0.00
7 S	Phenol-d6	80.000	76.395	4.5	85	0.00
8	2-Chlorophenol	40.000	38.454	3.9	87	0.00
9	Benzaldehyde	40.000	35.916	10.2	92	0.00
10 C	Phenol	40.000	38.898	2.8	85	0.00
11	bis(2-Chloroethyl)ether	40.000	38.134	4.7	85	0.00
12	1,3-Dichlorobenzene	40.000	38.473	3.8	90	0.00
13 C	1,4-Dichlorobenzene	40.000	38.282	4.3	89	0.00
14	1,2-Dichlorobenzene	40.000	38.609	3.5	88	0.00
15	Benzyl Alcohol	40.000	39.336	1.7	85	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.561	3.6	87	0.00
17	2-Methylphenol	40.000	38.608	3.5	86	0.00
18	Hexachloroethane	40.000	38.683	3.3	89	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.784	3.0	86	0.00
20	3+4-Methylphenols	40.000	38.504	3.7	85	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	88	0.00
22	Acetophenone	40.000	38.708	3.2	86	0.00
23 S	Nitrobenzene-d5	80.000	77.913	2.6	86	0.00
24	Nitrobenzene	40.000	38.653	3.4	86	0.00
25	Isophorone	40.000	39.257	1.9	85	0.00
26 C	2-Nitrophenol	40.000	39.419	1.5	86	0.00
27	2,4-Dimethylphenol	40.000	40.856	-2.1	86	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.829	0.4	88	0.00
29 C	2,4-Dichlorophenol	40.000	39.620	1.0	88	0.00
30	1,2,4-Trichlorobenzene	40.000	39.557	1.1	89	0.00
31	Naphthalene	40.000	39.229	1.9	88	0.00
32	Benzoic acid	40.000	40.607	-1.5	91	0.00
33	4-Chloroaniline	40.000	39.114	2.2	82	0.00
34 C	Hexachlorobutadiene	40.000	40.199	-0.5	90	0.00
35	Caprolactam	40.000	39.703	0.7	85	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.773	0.6	85	0.00
37	2-Methylnaphthalene	40.000	39.693	0.8	88	0.00
38	1-Methylnaphthalene	40.000	39.616	1.0	88	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	88	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.139	2.2	90	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.607	-1.5	93	0.00
42 S	2,4,6-Tribromophenol	80.000	80.016	-0.0	87	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.701	0.7	88	0.00
44	2,4,5-Trichlorophenol	40.000	39.898	0.3	87	0.00
45 S	2-Fluorobiphenyl	80.000	78.452	1.9	90	0.00
46	1,1'-Biphenyl	40.000	39.120	2.2	89	0.00
47	2-Chloronaphthalene	40.000	39.066	2.3	88	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112224\
 Data File : BF140564.D
 Acq On : 22 Nov 2024 09:20
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 22 10:52:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.155	2.1	85	0.00
49	Acenaphthylene	40.000	39.007	2.5	88	0.00
50	Dimethylphthalate	40.000	39.381	1.5	88	0.00
51	2,6-Dinitrotoluene	40.000	39.500	1.3	86	0.00
52 C	Acenaphthene	40.000	39.108	2.2	88	0.00
53	3-Nitroaniline	40.000	39.187	2.0	84	0.00
54 P	2,4-Dinitrophenol	40.000	40.859	-2.1	89	0.00
55	Dibenzofuran	40.000	38.827	2.9	87	0.00
56 P	4-Nitrophenol	40.000	39.134	2.2	83	0.00
57	2,4-Dinitrotoluene	40.000	39.837	0.4	86	0.00
58	Fluorene	40.000	38.881	2.8	88	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.303	-5.8	91	0.00
60	Diethylphthalate	40.000	39.734	0.7	89	0.00
61	4-Chlorophenyl-phenylether	40.000	39.607	1.0	89	0.00
62	4-Nitroaniline	40.000	39.279	1.8	85	0.00
63	Azobenzene	40.000	39.132	2.2	88	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	88	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.482	-11.2	91	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.092	2.3	88	0.00
67	4-Bromophenyl-phenylether	40.000	40.907	-2.3	93	0.00
68	Hexachlorobenzene	40.000	39.966	0.1	90	0.00
69	Atrazine	40.000	40.913	-2.3	107	0.00
70 C	Pentachlorophenol	40.000	44.717	-11.8	89	0.00
71	Phenanthrene	40.000	38.552	3.6	86	0.00
72	Anthracene	40.000	39.532	1.2	88	0.00
73	Carbazole	40.000	39.326	1.7	88	0.00
74	Di-n-butylphthalate	40.000	41.568	-3.9	93	0.00
75 C	Fluoranthene	40.000	40.287	-0.7	91	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	102	0.00
77	Benzydine	40.000	40.202	-0.5	103	0.00
78	Pyrene	40.000	36.647	8.4	93	0.00
79 S	Terphenyl-d14	80.000	75.119	6.1	96	0.00
80	Butylbenzylphthalate	40.000	40.936	-2.3	102	0.00
81	Benzo(a)anthracene	40.000	38.280	4.3	100	0.00
82	3,3'-Dichlorobenzidine	40.000	38.855	2.9	95	0.00
83	Chrysene	40.000	39.289	1.8	101	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.277	-10.7	114	0.00
85 c	Di-n-octyl phthalate	40.000	47.227	-18.1	122	0.00
86 I	Perylene-d12	20.000	20.000	0.0	87	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.657	8.4	80	0.00
88	Benzo(b)fluoranthene	40.000	44.420	-11.1	102	0.00
89	Benzo(k)fluoranthene	40.000	39.286	1.8	85	0.00
90 C	Benzo(a)pyrene	40.000	40.097	-0.2	88	0.00
91	Dibenzo(a,h)anthracene	40.000	37.672	5.8	82	0.00
92	Benzo(g,h,i)perylene	40.000	36.409	9.0	79	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112224\
Data File : BF140564.D
Acq On : 22 Nov 2024 09:20
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 22 10:52:08 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 21 15:23:48 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0