

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111324

Review By	yogesh	Review On	11/14/2024 3:08:53 AM
Supervise By	mohammad	Supervise On	11/15/2024 5:49:17 AM
SubDirectory	BF111324	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6573 SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6624 S12327,10ul/1000ul sample SP6559

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140331.D	13 Nov 2024 08:35		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF140332.D	13 Nov 2024 09:01		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF140333.D	13 Nov 2024 09:27	Compound#41,54,77 removed from 5 ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF140334.D	13 Nov 2024 09:53		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BF140335.D	13 Nov 2024 10:29		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF140336.D	13 Nov 2024 10:55	Injection Error	RC/JU	Not Ok
7	SSTDICC050	SSTDICC050	BF140337.D	13 Nov 2024 11:21		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF140338.D	13 Nov 2024 11:47		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF140339.D	13 Nov 2024 12:13	Compound#09 removed from 80 ppm	RC/JU	Ok,M
10	SSTDICCC040	SSTDICCC040	BF140340.D	13 Nov 2024 12:48		RC/JU	Ok
11	SSTDICV040	ICVBF111324	BF140341.D	13 Nov 2024 13:19		RC/JU	Ok
12	PB164401BL	PB164401BL	BF140342.D	13 Nov 2024 13:45		RC/JU	Ok
13	P4495-11	PT-BNA-SOIL	BF140343.D	13 Nov 2024 15:25	PT Sample, Need 5X Dilution	RC/JU	Dilution
14	P4495-11DL	PT-BNA-SOILDL	BF140344.D	13 Nov 2024 16:02		RC/JU	Ok
15	P3845-03	PT-BN-WP	BF140345.D	13 Nov 2024 16:40	PT Sample Analyzed for confirmation - Need 5X Dilution.	RC/JU	Dilution
16	P3845-03DL	PT-BN-WPDL	BF140346.D	13 Nov 2024 17:06		RC/JU	Ok
17	P3845-06	PT-ACIDS-WP	BF140347.D	13 Nov 2024 17:32	PT Sample Analyzed for confirmation - Need 5X Dilution.	RC/JU	Dilution

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111324

Review By	yogesh	Review On	11/14/2024 3:08:53 AM
Supervise By	mohammad	Supervise On	11/15/2024 5:49:17 AM
SubDirectory	BF111324	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	P3845-06DL	PT-ACIDS-WPDL	BF140348.D	13 Nov 2024 17:58	Dilution not matching (Low results)	RC/JU	Not Ok
19	SP6681	SP6681	BF140349.D	13 Nov 2024 18:39	8270-625.1 SPIKE SOLUTION	RC/JU	Ok,NR

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111624

Review By	yogesh	Review On	11/19/2024 3:19:51 AM
Supervise By	mohammad	Supervise On	11/19/2024 3:30:32 AM
SubDirectory	BF111624	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140416.D	16 Nov 2024 09:08		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140417.D	16 Nov 2024 09:34		RC/JU	Ok,M
3	PB164993BL	PB164993BL	BF140418.D	16 Nov 2024 10:02		RC/JU	Ok
4	PB164993BS	PB164993BS	BF140419.D	16 Nov 2024 10:28		RC/JU	Ok,M
5	P4821-04	BP-F-24	BF140420.D	16 Nov 2024 11:01	Internal Standard and Surrogate Fail	RC/JU	ReRun
6	P4821-08	BP-F-2	BF140421.D	16 Nov 2024 11:27		RC/JU	Ok
7	P4823-04	MH-4	BF140422.D	16 Nov 2024 11:53		RC/JU	Ok
8	P4833-04	MH-731	BF140423.D	16 Nov 2024 12:19		RC/JU	Ok
9	P4833-08	TP-2	BF140424.D	16 Nov 2024 12:46		RC/JU	Ok
10	P4824-01	T-20	BF140425.D	16 Nov 2024 13:12	Internal standard fail - hydrocarbons	RC/JU	Ok
11	P4821-05	BP-F-2	BF140426.D	16 Nov 2024 13:38		RC/JU	Ok
12	P4833-01	MH-731	BF140427.D	16 Nov 2024 14:04		RC/JU	Ok
13	P4833-01MS	MH-731MS	BF140428.D	16 Nov 2024 14:31		RC/JU	Ok,M
14	P4833-01MSD	MH-731MSD	BF140429.D	16 Nov 2024 14:57		RC/JU	Ok,M
15	P4848-01	TP-1	BF140430.D	16 Nov 2024 15:23		RC/JU	Ok,M
16	P4821-01	BP-F-24	BF140431.D	16 Nov 2024 15:49		RC/JU	Ok
17	P4821-01MS	BP-F-24MS	BF140432.D	16 Nov 2024 16:16		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111624

Review By	yogesh	Review On	11/19/2024 3:19:51 AM		
Supervise By	mohammad	Supervise On	11/19/2024 3:30:32 AM		
SubDirectory	BF111624	HP Acquire Method	BNA_F	HP Processing Method	Bf111324
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621			
CCC		SP6624			
Internal Standard/PEM		S12327,10ul/1000ul sample			
ICV/I.BLK		SP6559			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	P4821-01MSD	BP-F-24MSD	BF140433.D	16 Nov 2024 16:42		RC/JU	Ok,M
19	P4823-01	MH-4	BF140434.D	16 Nov 2024 17:08		RC/JU	Ok,M
20	P4768-02	B-3-2	BF140435.D	16 Nov 2024 17:34		RC/JU	Ok,M
21	P4833-05	TP-2	BF140436.D	16 Nov 2024 18:00		RC/JU	Ok
22	P4849-01	RR-1	BF140437.D	16 Nov 2024 18:27	Internal Standard Fail	RC/JU	ReRun
23	P4768-04	B-4	BF140438.D	16 Nov 2024 18:53		RC/JU	Ok
24	P4768-03	B-3-3	BF140439.D	16 Nov 2024 19:19	Internal Standard Fail	RC/JU	ReRun
25	P4859-01	NB-08-111424	BF140440.D	16 Nov 2024 19:45	Internal Standard Fail	RC/JU	ReRun
26	P4857-01	ETGI-328	BF140441.D	16 Nov 2024 20:11	Internal standard fail - hydrocarbons and oily matrix	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111824

Review By	yogesh	Review On	11/19/2024 3:20:55 AM
Supervise By	mohammad	Supervise On	11/19/2024 3:30:54 AM
SubDirectory	BF111824	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140442.D	18 Nov 2024 09:18		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140443.D	18 Nov 2024 10:13		RC/JU	Ok,M
3	PB165029BL	PB165029BL	BF140444.D	18 Nov 2024 10:39		RC/JU	Ok
4	PB165029BS	PB165029BS	BF140445.D	18 Nov 2024 11:05		RC/JU	Ok,M
5	PB164846BS	PB164846BS	BF140446.D	18 Nov 2024 11:32		RC/JU	Ok,M
6	PB164846BSD	PB164846BSD	BF140447.D	18 Nov 2024 11:58		RC/JU	Ok,M
7	PB164940BL	PB164940BL	BF140448.D	18 Nov 2024 12:24		RC/JU	Ok
8	PB164940BS	PB164940BS	BF140449.D	18 Nov 2024 12:50		RC/JU	Ok,M
9	PB164965BL	PB164965BL	BF140450.D	18 Nov 2024 13:16		RC/JU	Ok
10	PB164965BS	PB164965BS	BF140451.D	18 Nov 2024 13:43		RC/JU	Ok,M
11	PB164969BS	PB164969BS	BF140452.D	18 Nov 2024 14:19		RC/JU	Ok,M
12	PB164969BL	PB164969BL	BF140453.D	18 Nov 2024 14:45		RC/JU	Ok
13	PB164880TB	PB164880TB	BF140454.D	18 Nov 2024 15:11		RC/JU	Ok
14	DFTPP	DFTPP	BF140455.D	18 Nov 2024 15:48		RC/JU	Ok
15	SSTDCCC040	SSTDCCC040	BF140456.D	18 Nov 2024 16:14		RC/JU	Ok,M
16	PB164944TB	PB164944TB	BF140457.D	18 Nov 2024 16:40		RC/JU	Ok
17	P4821-04RE	BP-F-24RE	BF140458.D	18 Nov 2024 17:13	Surrogate Fail	RC/JU	Confirms
18	P4861-01	WC-11-A-202411	BF140459.D	18 Nov 2024 17:40		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111824

Review By	yogesh	Review On	11/19/2024 3:20:55 AM
Supervise By	mohammad	Supervise On	11/19/2024 3:30:54 AM
SubDirectory	BF111824	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12327,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	P4871-01	WC-10-A-202411	BF140460.D	18 Nov 2024 18:06		RC/JU	Ok
20	P4848-02	TP-1	BF140461.D	18 Nov 2024 18:32		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112024

Review By	yogesh	Review On	11/22/2024 3:51:12 AM				
Supervise By	mohammad	Supervise On	11/22/2024 4:00:42 AM				
SubDirectory	BF112024	HP Acquire Method	BNA_F	HP Processing Method	Bf111324		
STD. NAME		STD REF.#					
Tune/Reschk		SP6573					
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621					
CCC		SP6624					
Internal Standard/PEM		S12329,10ul/1000ul sample					
ICV/I.BLK		SP6559					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140487.D	20 Nov 2024 09:04		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140488.D	20 Nov 2024 09:31		RC/JU	Ok,M
3	PB165039BL	PB165039BL	BF140489.D	20 Nov 2024 09:57		RC/JU	Ok
4	PB165039BS	PB165039BS	BF140490.D	20 Nov 2024 10:23		RC/JU	Ok,M
5	PB165039BSD	PB165039BSD	BF140491.D	20 Nov 2024 10:49		RC/JU	Ok,M
6	P4843-01	SW-WTS-01	BF140492.D	20 Nov 2024 11:23		RC/JU	Ok,M
7	P4868-01	RW5-SP100-20241114	BF140493.D	20 Nov 2024 11:49		RC/JU	Ok
8	P4868-03	RW5-SP303-20241114	BF140494.D	20 Nov 2024 12:15	Internal Standard Fail	RC/JU	ReRun
9	P4892-02	WB-310-BOT	BF140495.D	20 Nov 2024 12:42		RC/JU	Ok
10	P4892-02MS	WB-310-BOTMS	BF140496.D	20 Nov 2024 13:08		RC/JU	Ok,M
11	P4892-02MSD	WB-310-BOTMSD	BF140497.D	20 Nov 2024 13:35		RC/JU	Ok,M
12	P4868-03RE	RW5-SP303-20241114	BF140498.D	20 Nov 2024 14:01	Internal Standard Fail	RC/JU	Confirms
13	SSTDCCC040	SSTDCCC040EC	BF140499.D	20 Nov 2024 14:34		RC/JU	Ok,M
14	DFTPP	DFTPP	BF140500.D	20 Nov 2024 15:33		RC/JU	Ok
15	SSTDCCC040	SSTDCCC040	BF140501.D	20 Nov 2024 15:59		RC/JU	Ok,M
16	PB165086BL	PB165086BL	BF140502.D	20 Nov 2024 16:25		RC/JU	Ok
17	P4849-02	RR-1	BF140503.D	20 Nov 2024 16:58		RC/JU	Ok
18	P4870-07	MH-736	BF140504.D	20 Nov 2024 17:24		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112024

Review By	yogesh	Review On	11/22/2024 3:51:12 AM		
Supervise By	mohammad	Supervise On	11/22/2024 4:00:42 AM		
SubDirectory	BF112024	HP Acquire Method	BNA_F	HP Processing Method	Bf111324
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621			
CCC		SP6624			
Internal Standard/PEM		S12329,10ul/1000ul sample			
ICV/I.BLK		SP6559			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P4910-01	MH-COTTAGE	BF140505.D	20 Nov 2024 17:51		RC/JU	Ok
20	P4893-01	MH-763	BF140506.D	20 Nov 2024 18:17		RC/JU	Ok
21	P4893-05	MH-762	BF140507.D	20 Nov 2024 18:43		RC/JU	Ok
22	P4768-05	B-5	BF140508.D	20 Nov 2024 19:09	Internal Standard Fail	RC/JU	Ok
23	P4870-10	TP-15	BF140509.D	20 Nov 2024 19:35		RC/JU	Ok
24	P4892-01	WB-310-TOP	BF140510.D	20 Nov 2024 20:02		RC/JU	Ok,M
25	P4909-01	BU-02-111824	BF140511.D	20 Nov 2024 20:28		RC/JU	Ok,M
26	P4910-05	MH-759	BF140512.D	20 Nov 2024 20:54	Internal Standard Fail	RC/JU	Ok,M
27	P4889-05DL	#72-11930DL	BF140513.D	20 Nov 2024 21:21	Internal Standard Fail	RC/JU	Ok,M
28	P4822-10	SOIL-5	BF140514.D	20 Nov 2024 21:47		RC/JU	Ok,M
29	P4822-12	SOIL-6	BF140515.D	20 Nov 2024 22:13		RC/JU	Ok,M
30	P4822-06	SOIL-3	BF140516.D	20 Nov 2024 22:40		RC/JU	Ok,M
31	P4822-08	SOIL-4	BF140517.D	20 Nov 2024 23:06		RC/JU	Ok,M
32	P4822-02	SOIL-1	BF140518.D	20 Nov 2024 23:32	Internal Standard Fail	RC/JU	ReRun
33	P4822-04	SOIL-2	BF140519.D	20 Nov 2024 23:58	Internal Standard Fail	RC/JU	ReRun
34	P4887-05	MH-760	BF140520.D	21 Nov 2024 00:24	Internal Standard Fail	RC/JU	ReRun
35	P4860-02	PH2-BOT-001	BF140521.D	21 Nov 2024 00:51	Internal Standard Fail	RC/JU	ReRun
36	P4860-03	PH2-BOT-002	BF140522.D	21 Nov 2024 01:17	Internal Standard Fail	RC/JU	ReRun
37	P4860-07	PH2-BOT-008	BF140523.D	21 Nov 2024 01:43	Internal Standard Fail	RC/JU	ReRun

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112024

Review By	yogesh	Review On	11/22/2024 3:51:12 AM		
Supervise By	mohammad	Supervise On	11/22/2024 4:00:42 AM		
SubDirectory	BF112024	HP Acquire Method	BNA_F	HP Processing Method	Bf111324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12329,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

38	P4916-01	TP-1-WC	BF140524.D	21 Nov 2024 02:09	Internal Standard Fail	RC/JU	ReRun
39	P4916-05	TP-2-WC	BF140525.D	21 Nov 2024 02:35	Internal Standard Fail	RC/JU	ReRun

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112124

Review By	yogesh	Review On	11/22/2024 3:53:36 AM
Supervise By	mohammad	Supervise On	11/22/2024 4:01:03 AM
SubDirectory	BF112124	HP Acquire Method	BNA_F
		HP Processing Method	bf112124

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12329,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140526.D	21 Nov 2024 10:17		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140527.D	21 Nov 2024 10:44	A Fresh Calibration is required.	RC/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BF140528.D	21 Nov 2024 11:13		RC/JU	Ok
4	SSTDICC005	SSTDICC005	BF140529.D	21 Nov 2024 11:39	Compound#9,32,41,54,56,65,70 removed from 5 ppm	RC/JU	Ok
5	SSTDICC010	SSTDICC010	BF140530.D	21 Nov 2024 12:05		RC/JU	Ok,M
6	SSTDICC020	SSTDICC020	BF140531.D	21 Nov 2024 12:32	Compound#32,41,54 Kept on LR	RC/JU	Ok,M
7	SSTDICCC040	SSTDICCC040	BF140532.D	21 Nov 2024 12:58	Calibration failed for Benzidine	RC/JU	Ok,M
8	SSTDICC050	SSTDICC050	BF140533.D	21 Nov 2024 13:25	The Calibration is Good For 8270 DOD Except com#77 and good for 625.1 Method Except com#77	RC/JU	Ok
9	SSTDICC060	SSTDICC060	BF140534.D	21 Nov 2024 13:51		RC/JU	Ok
10	SSTDICC080	SSTDICC080	BF140535.D	21 Nov 2024 14:18	Compound#9 removed from 80 ppm	RC/JU	Ok
11	SSTDICV040	ICVBF112124	BF140536.D	21 Nov 2024 15:07		RC/JU	Ok,M
12	PB165144BL	PB165144BL	BF140537.D	21 Nov 2024 15:34		RC/JU	Ok
13	DFTPP	DFTPP	BF140538.D	21 Nov 2024 16:27		RC/JU	Ok
14	SSTDCCC040	SSTDCCC040	BF140539.D	21 Nov 2024 16:54		RC/JU	Ok,M
15	PB165060TB	PB165060TB	BF140540.D	21 Nov 2024 17:20		RC/JU	Ok
16	P4887-06	MH-760	BF140541.D	21 Nov 2024 17:55		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112124

Review By	yogesh	Review On	11/22/2024 3:53:36 AM		
Supervise By	mohammad	Supervise On	11/22/2024 4:01:03 AM		
SubDirectory	BF112124	HP Acquire Method	BNA_F	HP Processing Method	bf112124
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621			
CCC		SP6624			
Internal Standard/PEM		S12329,10ul/1000ul sample			
ICV/I.BLK		SP6559			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

17	P4887-02	MH-739	BF140542.D	21 Nov 2024 18:21		RC/JU	Ok
18	P4870-16	TP-15	BF140543.D	21 Nov 2024 18:48		RC/JU	Ok
19	P4870-15	MH-736	BF140544.D	21 Nov 2024 19:14		RC/JU	Ok
20	P4870-14	MH-735	BF140545.D	21 Nov 2024 19:40		RC/JU	Ok
21	P4870-13	TP-1	BF140546.D	21 Nov 2024 20:07		RC/JU	Ok
22	P4929-02	ARS520	BF140547.D	21 Nov 2024 20:33		RC/JU	Ok
23	P4924-04	MH-4	BF140548.D	21 Nov 2024 20:59		RC/JU	Ok
24	P4921-01	WC-11-A-202411	BF140549.D	21 Nov 2024 21:25		RC/JU	Ok
25	P4916-12	TP-3-WC	BF140550.D	21 Nov 2024 21:51		RC/JU	Ok
26	P4916-08	TP-2-WC	BF140551.D	21 Nov 2024 22:18		RC/JU	Ok
27	P4916-04	TP-1-WC	BF140552.D	21 Nov 2024 22:44		RC/JU	Ok
28	P4887-01	MH-739	BF140553.D	21 Nov 2024 23:10		RC/JU	Ok
29	P4916-09	TP-3-WC	BF140554.D	21 Nov 2024 23:36		RC/JU	Ok
30	P4916-09MS	TP-3-WCMS	BF140555.D	22 Nov 2024 00:02		RC/JU	Ok,NR
31	P4916-09MSD	TP-3-WCMSD	BF140556.D	22 Nov 2024 00:28		RC/JU	Ok,NR
32	P4916-01	TP-1-WC	BF140557.D	22 Nov 2024 00:54		RC/JU	Ok
33	P4916-05RE	TP-2-WCRE	BF140558.D	22 Nov 2024 01:21	Internal Standard Fail	RC/JU	Confirms
34	P4887-05RE	MH-760RE	BF140559.D	22 Nov 2024 01:47	Internal Standard Fail	RC/JU	Confirms
35	P4924-01	MH-4	BF140560.D	22 Nov 2024 02:13	Internal Standard Fail	RC/JU	ReRun

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112124

Review By	yogesh	Review On	11/22/2024 3:53:36 AM		
Supervise By	mohammad	Supervise On	11/22/2024 4:01:03 AM		
SubDirectory	BF112124	HP Acquire Method	BNA_F	HP Processing Method	bf112124

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12329,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

36	P4936-01	PL-01-11202024	BF140561.D	22 Nov 2024 02:39	Internal Standard Fail	RC/JU	Ok,NR
37	P4929-01	ARS520	BF140562.D	22 Nov 2024 03:06	Internal Standard Fail	RC/JU	Ok,NR

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112224

Review By		Review On			
Supervise By		Supervise On			
SubDirectory	BF112224	HP Acquire Method	BNA_F	HP Processing Method	bf112124
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621			
CCC		SP6624			
Internal Standard/PEM		S12329,10ul/1000ul sample			
ICV/I.BLK		SP6559			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140563.D	22 Nov 2024 08:54		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140564.D	22 Nov 2024 09:20		RC/JU	Ok
3	PB165129BL	PB165129BL	BF140565.D	22 Nov 2024 09:46		RC/JU	Ok
4	PB165129BS	PB165129BS	BF140566.D	22 Nov 2024 10:12		RC/JU	Ok,NR
5	P4910-08	MH-759	BF140567.D	22 Nov 2024 10:44		RC/JU	Ok
6	P4910-04	MH-COTTAGE	BF140568.D	22 Nov 2024 11:10		RC/JU	Ok
7	P4893-08	MH-762	BF140569.D	22 Nov 2024 11:36		RC/JU	Ok
8	P4893-04	MH-763	BF140570.D	22 Nov 2024 12:02		RC/JU	Ok
9	P4925-08	MH-741	BF140571.D	22 Nov 2024 12:28		RC/JU	Ok
10	P4925-04	MH-741	BF140572.D	22 Nov 2024 12:55		RC/JU	Ok
11	P4923-06	72-11991	BF140573.D	22 Nov 2024 13:21		RC/JU	Ok
12	P4925-01	MH-741	BF140574.D	22 Nov 2024 13:47		RC/JU	Ok
13	P4938-01	MH-732	BF140575.D	22 Nov 2024 14:13		RC/JU	Ok
14	P4938-01MS	MH-732MS	BF140576.D	22 Nov 2024 14:40		RC/JU	Ok,NR
15	P4938-01MSD	MH-732MSD	BF140577.D	22 Nov 2024 15:06		RC/JU	Ok,NR
16	P4860-08	PH2-BOT-007	BF140578.D	22 Nov 2024 15:32		RC/JU	Ok
17	P4860-05	PH2-BOT-004	BF140579.D	22 Nov 2024 15:58		RC/JU	Ok
18	P4924-01	MH-4	BF140580.D	22 Nov 2024 16:24		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF112224

Review By		Review On			
Supervise By		Supervise On			
SubDirectory	BF112224	HP Acquire Method	BNA_F	HP Processing Method	bf112124
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12329,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P4948-01	337	BF140581.D	22 Nov 2024 16:50		RC/JU	Ok,NR
20	P4949-03	CONTAINMENT-STON	BF140582.D	22 Nov 2024 17:16	Internal Standard Fail	RC/JU	ReRun
21	P4948-03	72-11944	BF140583.D	22 Nov 2024 17:43	Internal Standard Fail	RC/JU	ReRun
22	P4822-04	SOIL-2	BF140584.D	22 Nov 2024 18:09		RC/JU	Ok
23	P4822-02	SOIL-1	BF140585.D	22 Nov 2024 18:35		RC/JU	Ok,NR
24	P4925-05	MH-758	BF140586.D	22 Nov 2024 19:01	Typo corrected from P4925-04 to P4925-05,Please confirm	RC/JU	Ok
25	P4938-05	MH-734	BF140587.D	22 Nov 2024 19:27		RC/JU	Ok
26	P4860-02	PH2-BOT-001	BF140588.D	22 Nov 2024 19:53		RC/JU	Ok

M : Manual Integration