

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/18/2024

Lab File ID: BF140443.D Time Analyzed: 10:13

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	140634	6.875	521097	8.16	286394	9.92
UPPER LIMIT	281268	7.375	1042190	8.657	572788	10.416
LOWER LIMIT	70317	6.375	260549	7.657	143197	9.416
EPA SAMPLE NO.						
01 PB164940BL	137693	6.88	537124	8.15	302062	9.91
02 PB164940BS	140752	6.88	537412	8.16	299132	9.91

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.