

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/18/2024

Lab File ID: BF140443.D Time Analyzed: 10:13

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	525142	11.404	270046	14.074	249102	15.604
UPPER LIMIT	1050280	11.904	540092	14.574	498204	16.104
LOWER LIMIT	262571	10.904	135023	13.574	124551	15.104
EPA SAMPLE NO.						
01 PB164940BL	571216	11.40	335934	14.05	283190	15.56
02 PB164940BS	551457	11.40	299750	14.06	260212	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.