

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: BAPS03

Lab Code: CHEM

SAS No.: P4822 SDG NO.: P4822

Lab File ID: BF140416.D

DFTPP Injection Date: 11/16/2024

Instrument ID: BNA_F

DFTPP Injection Time: 09:08

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	37.4
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	37.8
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	49
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	28.4
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	14.9
442	Greater than 50% of mass 198	93.1
443	15.0 - 24.0% of mass 442	17.5 (18.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140417.D	11/16/2024	09:34
BP-F-24MS	P4821-01MS	BF140432.D	11/16/2024	16:16
BP-F-24MSD	P4821-01MSD	BF140433.D	11/16/2024	16:42