

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: BAPS03
Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG No.: P4822
Instrument ID: MSVOA_Y Calibration Date(s): 10/30/2024 10/30/2024
Heated Purge: (Y/N) Y Calibration Time(s): 12:39 14:52
GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF005 = VY020075.D	RRF010 = VY020076.D	RRF020 = VY020077.D	RRF050 = VY020078.D	RRF100 = VY020079.D	RRF150 = VY020080.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.507	0.488	0.482	0.381	0.456	0.463	0.463	9.5
Chloromethane	0.831	0.835	0.791	0.683	0.731	0.728	0.767	8.1
Vinyl Chloride	0.867	0.874	0.835	0.754	0.809	0.799	0.823	5.5
Bromomethane	0.631	0.574	0.545	0.480	0.521	0.528	0.546	9.5
Chloroethane	0.562	0.555	0.546	0.486	0.527	0.511	0.531	5.5
Trichlorofluoromethane	1.025	1.018	1.037	0.927	1.012	1.014	1.005	3.9
1,1,2-Trichlorotrifluoroethane	0.597	0.584	0.567	0.509	0.561	0.565	0.564	5.3
1,1-Dichloroethene	0.534	0.578	0.536	0.493	0.543	0.552	0.540	5.1
Acetone	0.134	0.138	0.138	0.140	0.165	0.155	0.145	8.5
Carbon Disulfide	1.790	1.835	1.806	1.674	1.832	1.837	1.796	3.5
Methyl tert-butyl Ether	1.229	1.344	1.376	1.241	1.483	1.471	1.357	8
Methyl Acetate	0.314	0.357	0.351	0.317	0.362	0.342	0.340	6
Methylene Chloride	0.807	0.716	0.641	0.555	0.597	0.597	0.652	14.3
trans-1,2-Dichloroethene	0.611	0.618	0.613	0.564	0.616	0.628	0.609	3.7
1,1-Dichloroethane	1.155	1.209	1.189	1.070	1.194	1.200	1.170	4.5
Cyclohexane	1.284	1.206	1.138	0.986	1.093	1.088	1.132	9.1
2-Butanone	0.187	0.197	0.192	0.179	0.219	0.206	0.197	7.3
Carbon Tetrachloride	0.467	0.491	0.499	0.462	0.519	0.524	0.494	5.2
cis-1,2-Dichloroethene	0.646	0.714	0.715	0.646	0.740	0.736	0.700	6.1
Bromochloromethane	0.572	0.562	0.527	0.505	0.553	0.567	0.548	4.8
Chloroform	1.141	1.170	1.186	1.068	1.196	1.203	1.161	4.3
1,1,1-Trichloroethane	0.986	1.026	1.009	0.950	1.041	1.069	1.013	4.1
Methylcyclohexane	0.603	0.602	0.631	0.578	0.665	0.661	0.623	5.6
Benzene	1.414	1.405	1.455	1.325	1.504	1.498	1.433	4.7
1,2-Dichloroethane	0.379	0.391	0.402	0.365	0.425	0.418	0.397	5.8
Trichloroethene	0.328	0.331	0.332	0.307	0.348	0.346	0.332	4.4
1,2-Dichloropropane	0.323	0.331	0.357	0.320	0.363	0.364	0.343	6
Bromodichloromethane	0.459	0.476	0.487	0.451	0.526	0.522	0.487	6.4
4-Methyl-2-Pentanone	0.190	0.223	0.238	0.219	0.269	0.252	0.232	11.9
Toluene	0.803	0.859	0.900	0.833	0.953	0.953	0.884	7.1

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: BAPS03
 Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG No.: P4822
 Instrument ID: MSVOA_Y Calibration Date(s): 10/30/2024 10/30/2024
 Heated Purge: (Y/N) Y Calibration Time(s): 12:39 14:52
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY020075.D		RRF010 = VY020076.D		RRF020 = VY020077.D			
		RRF050 = VY020078.D		RRF100 = VY020079.D		RRF150 = VY020080.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
t-1,3-Dichloropropene	0.369	0.400	0.436	0.401	0.486	0.487	0.430	11.3	
cis-1,3-Dichloropropene	0.446	0.485	0.518	0.484	0.573	0.573	0.513	10	
1,1,2-Trichloroethane	0.206	0.225	0.235	0.216	0.257	0.248	0.231	8.4	
2-Hexanone	0.131	0.161	0.168	0.158	0.198	0.182	0.166	13.7	
Dibromochloromethane	0.256	0.281	0.300	0.273	0.332	0.326	0.295	10.3	
1,2-Dibromoethane	0.203	0.214	0.215	0.196	0.233	0.228	0.215	6.7	
Tetrachloroethene	0.326	0.316	0.344	0.307	0.344	0.343	0.330	4.8	
Chlorobenzene	1.024	1.044	1.094	0.993	1.114	1.124	1.065	5	
Ethyl Benzene	1.886	1.934	2.015	1.876	2.114	2.124	1.992	5.5	
m/p-Xylenes	0.685	0.693	0.742	0.692	0.779	0.782	0.729	6.2	
o-Xylene	0.652	0.662	0.686	0.649	0.745	0.745	0.690	6.5	
Styrene	0.985	1.080	1.154	1.093	1.259	1.273	1.141	9.8	
Bromoform	0.151	0.162	0.173	0.165	0.202	0.195	0.175	11.4	
Isopropylbenzene	4.065	3.829	3.995	3.685	4.099	4.293	3.994	5.3	
1,1,2,2-Tetrachloroethane	0.716	0.707	0.698	0.624	0.727	0.719	0.699	5.4	
1,3-Dichlorobenzene	1.659	1.521	1.680	1.509	1.703	1.774	1.641	6.4	
1,4-Dichlorobenzene	1.680	1.604	1.653	1.487	1.666	1.716	1.634	5	
1,2-Dichlorobenzene	1.391	1.384	1.460	1.310	1.483	1.530	1.426	5.6	
1,2-Dibromo-3-Chloropropane	0.107	0.104	0.106	0.092	0.114	0.112	0.106	7.6	
1,2,4-Trichlorobenzene	0.652	0.708	0.784	0.693	0.865	0.895	0.766	12.8	
1,2,3-Trichlorobenzene	0.578	0.606	0.653	0.575	0.737	0.754	0.650	12.1	
1,2-Dichloroethane-d4	0.608	0.657	0.578	0.586	0.668	0.647	0.624	6.1	
Dibromofluoromethane	0.305	0.317	0.300	0.305	0.340	0.340	0.318	5.7	
Toluene-d8	1.187	1.251	1.187	1.224	1.380	1.364	1.265	6.8	
4-Bromofluorobenzene	0.371	0.402	0.376	0.399	0.463	0.454	0.411	9.5	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.