

Instrument ID: MSVOA_D

Daily Analysis Runlog For Sequence/QC Batch ID # VD111524

Review By	Mahesh Dadoda	Review On	11/18/2024 3:13:25 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/18/2024 3:17:20 PM
SubDirectory	VD111524	HP Acquire Method	HP Processing Method 82d111424s.m
STD. NAME	STD REF.#		
Tune/Reschk	VP131514		
Initial Calibration Stds	VP131516,VP131518,VP131520,VP131522,VP131524,VP131526		
CCC	VP131535,VP131536		
Internal Standard/PEM	VP128297		
ICV/I.BLK	VP131528		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VD080010.D	14 Nov 2024 17:14	Method failed for Comp. #05	RP/MD	Not Ok
2	VSTDIC005	VSTDIC005	VD080011.D	14 Nov 2024 18:02	%D failed for Comp. #16 in 05PPB	RP/MD	Not Ok
3	VSTDIC010	VSTDIC010	VD080012.D	14 Nov 2024 18:29		RP/MD	Not Ok
4	VSTDIC020	VSTDIC020	VD080013.D	14 Nov 2024 18:56		RP/MD	Not Ok
5	VSTDIC050	VSTDIC050	VD080014.D	14 Nov 2024 19:22	Comp. #10,16 are on Quadratic Regression	RP/MD	Not Ok
6	VSTDIC100	VSTDIC100	VD080015.D	14 Nov 2024 19:49		RP/MD	Not Ok
7	VSTDIC150	VSTDIC150	VD080016.D	14 Nov 2024 20:16		RP/MD	Not Ok
8	VSTDICV050	ICVVD111524	VD080017.D	14 Nov 2024 20:43	ICV Failed for comp. #13	RP/MD	Not Ok
9	VIBLK	VIBLK	VD080018.D	14 Nov 2024 21:36		RP/MD	Not Ok
10	BFB	BFB	VD080019.D	15 Nov 2024 09:14		RP/MD	Ok
11	VSTDIC005	VSTDIC005	VD080020.D	15 Nov 2024 10:04	QR @ comp #05	RP/MD	Ok,M
12	VSTDIC010	VSTDIC010	VD080021.D	15 Nov 2024 10:31	%D fail for comp #05 in 10PPB	RP/MD	Ok,M
13	VSTDIC020	VSTDIC020	VD080022.D	15 Nov 2024 10:58		RP/MD	Ok,M
14	VSTDIC050	VSTDIC050	VD080023.D	15 Nov 2024 11:24		RP/MD	Ok,M
15	VSTDIC100	VSTDIC100	VD080024.D	15 Nov 2024 11:51		RP/MD	Ok,M
16	VSTDIC150	VSTDIC150	VD080025.D	15 Nov 2024 12:18		RP/MD	Ok,M
17	VSTDICV050	ICVVD111524	VD080026.D	15 Nov 2024 15:44		RP/MD	Ok,M

Instrument ID: MSVOA_D

Daily Analysis Runlog For Sequence/QC Batch ID # VD111524

Review By	Maresh Dadoda	Review On	11/18/2024 3:13:25 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/18/2024 3:17:20 PM
SubDirectory	VD111524	HP Acquire Method	HP Processing Method 82d111424s.m
STD. NAME	STD REF.#		
Tune/Reschk	VP131514		
Initial Calibration Stds	VP131516,VP131518,VP131520,VP131522,VP131524,VP131526		
CCC	VP131535,VP131536		
Internal Standard/PEM	VP128297		
ICV/I.BLK	VP131528		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	VD1115SBL01	VBLK21	VD080027.D	15 Nov 2024 16:19		RP/MD	Ok
19	VD1115SBS01	VD1115SBS01	VD080028.D	15 Nov 2024 16:46		RP/MD	Ok,M
20	P4837-03	3TRK-STONE	VD080029.D	15 Nov 2024 17:13	vial-A	RP/MD	Ok
21	P4836-03	T3-STONE	VD080030.D	15 Nov 2024 17:40	ISTD Fail vial-A	RP/MD	ReRun
22	P4835-03	345-2-STONE	VD080031.D	15 Nov 2024 18:07	vial-A	RP/MD	Ok
23	P4834-03	SLP-1-STONE	VD080032.D	15 Nov 2024 18:34	Internal Standard and Surrogate Fail VIAL- A	RP/MD	ReRun
24	P4838-03	T1-STONE	VD080033.D	15 Nov 2024 19:01	ISTD Fail :Vial- A	RP/MD	ReRun
25	P4833-03	MH-731-VOC	VD080034.D	15 Nov 2024 19:28	Internal not spike; Surrogate fail Vial- A	RP/MD	Not Ok
26	VD1115SBSD01	VD1115SBSD01	VD080035.D	15 Nov 2024 19:55		RP/MD	Ok,M
27	VSTDCCC050	VSTD05021	VD080036.D	15 Nov 2024 20:22		RP/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_D

Daily Analysis Runlog For Sequence/QC Batch ID # VD111724

Review By	Mahesh Dadoda	Review On	11/18/2024 3:14:40 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/18/2024 3:17:59 PM
SubDirectory	VD111724	HP Acquire Method	HP Processing Method 82D111424s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131566
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131568,VP131569 VP128297

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VD080037.D	17 Nov 2024 11:29		RP/MD	Ok
2	VSTDCCC050	VSTDCCC050	VD080038.D	17 Nov 2024 13:28		RP/MD	Ok,M
3	VD1117SBL01	VD1117SBL01	VD080039.D	17 Nov 2024 13:55		RP/MD	Ok
4	VD1117SBS01	VD1117SBS01	VD080040.D	17 Nov 2024 14:25		RP/MD	Ok,M
5	VD1117SBSD01	VD1117SBSD01	VD080041.D	17 Nov 2024 14:52		RP/MD	Ok,M
6	P4810-02	VOC	VD080042.D	17 Nov 2024 15:19	Not purge vial-A	RP/MD	Not Ok
7	P4838-03	T1-STONE	VD080043.D	17 Nov 2024 15:46	vial-B	RP/MD	Ok
8	P4833-03	MH-731-VOC	VD080044.D	17 Nov 2024 16:13	vial-B	RP/MD	Ok
9	P4836-03RE	T3-STONERE	VD080045.D	17 Nov 2024 16:40	Internal Standard Fail vial-B	RP/MD	Confirms
10	P4834-03RE	SLP-1-STONERE	VD080046.D	17 Nov 2024 17:07	Internal Standard Fail ;Surrogate fail VIAL- B	RP/MD	Confirms
11	P4833-07	TP-2-VOC	VD080047.D	17 Nov 2024 17:34	Internal Standard Fail ;Surrogate fail VIAL- A	RP/MD	ReRun
12	P4824-01	T-20	VD080048.D	17 Nov 2024 18:01	Internal Standard Fail ;Surrogate fail VIAL- A	RP/MD	ReRun
13	P4825-01	STOCK-PILE	VD080049.D	17 Nov 2024 18:28	Internal Standard Fail ;Surrogate fail VIAL- A	RP/MD	ReRun
14	P4857-03	VNJ-226	VD080050.D	17 Nov 2024 18:55	Internal Standard Fail vial-A	RP/MD	ReRun
15	P4859-01	NB-08-111424	VD080051.D	17 Nov 2024 19:22	Internal Standard Fail vial-A	RP/MD	ReRun
16	P4822-05RE	SOIL-3RE	VD080052.D	17 Nov 2024 19:49	Internal Standard Fail vial-A	RP/MD	Confirms
17	P4822-07RE	SOIL-4RE	VD080053.D	17 Nov 2024 20:16	Internal Standard Fail vial-B	RP/MD	Confirms

Instrument ID: MSVOA_D

Daily Analysis Runlog For Sequence/QC Batch ID # VD111724

Review By	Maresh Dadoda	Review On	11/18/2024 3:14:40 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/18/2024 3:17:59 PM
SubDirectory	VD111724	HP Acquire Method	HP Processing Method 82D111424s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131566
CCC	VP131568,VP131569
Internal Standard/PEM	VP128297
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	P4822-01	SOIL-1	VD080054.D	17 Nov 2024 20:43	Internal Standard Fail vial-A	RP/MD	Not Ok
19	P4821-07	BP-F-2-VOC	VD080055.D	17 Nov 2024 21:10	vial-A	RP/MD	Ok
20	VSTDCCC050	VSTDCCC050EC	VD080056.D	17 Nov 2024 21:37		RP/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY103024

Review By	Maresh Dadoda	Review On	11/1/2024 5:33:31 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/1/2024 5:33:47 PM
SubDirectory	VY103024	HP Acquire Method	HP Processing Method 82y103024s.m

STD. NAME	STD REF.#
Tune/Reschk	VP131199
Initial Calibration Stds	VP131200,VP131201,VP131203,VP131204,VP131205,VP131206
CCC	
Internal Standard/PEM	VP128297
ICV/I.BLK	VP131207
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020074.D	30 Oct 2024 12:07		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY020075.D	30 Oct 2024 12:39	Method pass	SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY020076.D	30 Oct 2024 13:02		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY020077.D	30 Oct 2024 13:24		SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY020078.D	30 Oct 2024 14:06		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY020079.D	30 Oct 2024 14:29		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY020080.D	30 Oct 2024 14:52		SY/MD	Ok,M
8	VIBLK	VIBLK	VY020081.D	30 Oct 2024 15:15		SY/MD	Ok
9	VSTDICV050	ICVVY103024	VY020082.D	30 Oct 2024 15:47	fail icv	SY/MD	Not Ok

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY111324

Review By	Maresh Dadoda	Review On	11/14/2024 3:19:21 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/14/2024 3:21:28 PM
SubDirectory	VY111324	HP Acquire Method	HP Processing Method 82y103024s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131442
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131443,VP131444 VP128297

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY020275.D	13 Nov 2024 09:22		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY020276.D	13 Nov 2024 10:21	CCAL failed high for comp. #16	SY/MD	Ok,M
3	VY1113SBL01	VY1113SBL01	VY020277.D	13 Nov 2024 12:00		SY/MD	Ok
4	VY1113SBS01	VY1113SBS01	VY020278.D	13 Nov 2024 12:43		SY/MD	Ok,M
5	VY1113SBSD01	VY1113SBSD01	VY020279.D	13 Nov 2024 13:05		SY/MD	Ok,M
6	P4811-01RE	RB24008RE	VY020280.D	13 Nov 2024 14:05	vial-B Internal Standard and Surrogate Fail	SY/MD	Confirms
7	P4807-11	BF-F13-VOC	VY020281.D	13 Nov 2024 14:29	vial-B	SY/MD	Ok
8	P4792-03	OILY-STONE-DRUM	VY020282.D	13 Nov 2024 14:52	vial-B	SY/MD	Ok
9	P4810-02	VOC	VY020283.D	13 Nov 2024 15:15	vial-A Internal Standard Fail	SY/MD	Ok
10	P4807-03	TP-7-VOC	VY020284.D	13 Nov 2024 15:39	vial-B	SY/MD	Ok
11	P4821-03	BP-F-24-VOC	VY020285.D	13 Nov 2024 16:02	vial-B	SY/MD	Ok
12	P4821-07	BP-F-2-VOC	VY020286.D	13 Nov 2024 16:26	vial-A CCAL failed high for comp. #16 ;BSD failed high for comp. #52	SY/MD	ReRun
13	P4823-03	MH-4-VOC	VY020287.D	13 Nov 2024 16:49	vial-A	SY/MD	Ok
14	P4822-03	SOIL-2	VY020288.D	13 Nov 2024 17:12	vial-A	SY/MD	Ok
15	P4822-01	SOIL-1	VY020289.D	13 Nov 2024 17:36	vial-A	SY/MD	Ok
16	P4822-05	SOIL-3	VY020290.D	13 Nov 2024 17:59	vial-A Internal Standard Fail	SY/MD	ReRun
17	P4822-11	SOIL-6	VY020291.D	13 Nov 2024 18:23	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY111324

Review By	Maresh Dadoda	Review On	11/14/2024 3:19:21 PM
Supervise By	Semsettin Yesilyurt	Supervise On	11/14/2024 3:21:28 PM
SubDirectory	VY111324	HP Acquire Method	HP Processing Method 82y103024s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP131442		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131443,VP131444 VP128297		

18	P4822-09	SOIL-5	VY020292.D	13 Nov 2024 18:46	vial-A	SY/MD	Ok
19	P4822-07	SOIL-4	VY020293.D	13 Nov 2024 19:09	Internal Standard Fail vial-A	SY/MD	ReRun
20	VY1113SBS02	VY1113SBS02	VY020294.D	13 Nov 2024 19:32		SY/MD	Ok,M
21	VY1113SBSD02	VY1113SBSD02	VY020295.D	13 Nov 2024 19:55	BSD failed high for comp. #52,30	SY/MD	Ok,M
22	VSTDCCC050	VSTDCCC050EC	VY020296.D	13 Nov 2024 20:18		SY/MD	Ok,M

M : Manual Integration