

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: BAPS03
 Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822
 Lab File ID: VD080038.D Date Analyzed: 11/17/2024
 Instrument ID: MSVOA_D Time Analyzed: 13:28
 GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	158443	13.516				
UPPER LIMIT	316886	14.016				
LOWER LIMIT	79221.5	13.016				
EPA SAMPLE NO.						
SOIL-3RE	75449 *	13.52				
SOIL-4RE	66849 *	13.52				
VD1117SBL01	133574	13.52				
VD1117SBS01	159474	13.52				
VD1117SBSD01	169441	13.52				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.