

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: BAPS03
Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822
Lab File ID: VY020276.D Date Analyzed: 11/13/2024
Instrument ID: MSVOA_Y Time Analyzed: 10:21
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	165760	13.352				
UPPER LIMIT	331520	13.852				
LOWER LIMIT	82880	12.852				
EPA SAMPLE NO.						
SOIL-1	124141	13.35				
SOIL-2	152978	13.35				
SOIL-3	79473 *	13.35				
SOIL-4	81233 *	13.35				
SOIL-5	137784	13.35				
SOIL-6	124535	13.35				
VY1113SBL01	108368	13.35				
VY1113SBS02	186078	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.