

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: BAPS03
Lab Code: CHEM Case No.: P4822 SAS No.: P4822 SDG NO.: P4822
Lab File ID: VD080019.D BFB Injection Date: 11/15/2024
Instrument ID: MSVOA_D BFB Injection Time: 09:14
GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.1
75	30.0 - 60.0% of mass 95	50.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.0 (0.0) 1
174	50.0 - 100.0% of mass 95	90.2
175	5.0 - 9.0% of mass 174	7.3 (8.1) 1
176	95.0 - 101.0% of mass 174	87.4 (97) 1
177	5.0 - 9.0% of mass 176	5.9 (6.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VD080020.D	11/15/2024	10:04
VSTDIC010	VSTDIC010	VD080021.D	11/15/2024	10:31
VSTDIC020	VSTDIC020	VD080022.D	11/15/2024	10:58
VSTDIC050	VSTDIC050	VD080023.D	11/15/2024	11:24
VSTDIC100	VSTDIC100	VD080024.D	11/15/2024	11:51
VSTDIC150	VSTDIC150	VD080025.D	11/15/2024	12:18