


LB133633

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.4956e-004 slope

C= -3.3584e-002 y-intercept

Rho= 0.9998942

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/17
0.00	0.000	-0.013	-0.013	83	0.000	83	0				
0.05	0.050					442	0				
0.20	0.200	0.220	0.020	1017	0.0 %	1017					
2.50	2.500	2.560	0.060	10392	0.0 %	10392					
5.00	5.000	4.932	-0.068	19899	0.0 %	19899					
7.50	7.500	7.435	-0.065	29926	0.0 %	29926					
10.0	10.000	10.066	0.066	40468	0.0 %	40468					

LB133633

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	83	-		0 SFAM01.1	PPB	11/26/2024 13:52	Std
	0.2 S01	1017	-		0.2 SFAM01.1	PPB	11/26/2024 13:54	Std
	2.5 S02	10392	-		2.5 SFAM01.1	PPB	11/26/2024 13:57	Std
	5 S03	19899	-		5 SFAM01.1	PPB	11/26/2024 13:59	Std
	7.5 S04	29926	-		7.5 SFAM01.1	PPB	11/26/2024 14:01	Std
	10 S05	40468	-		10 SFAM01.1	PPB	11/26/2024 14:04	Std
ICV001	ICV001	17194	4.2574 -		SFAM01.1	PPB	11/26/2024 14:09	SMPL
ICB001	ICB001	-159	-0.0733 -		SFAM01.1	PPB	11/26/2024 14:12	SMPL
CCV055	CCV055	19836	4.9167 -		SFAM01.1	PPB	11/26/2024 14:14	SMPL
CCB055	CCB055	-179	-0.0783 -		SFAM01.1	PPB	11/26/2024 14:16	SMPL
PB165275BL	PBW275	127	-0.0019 -		SFAM01.1	PPB	11/26/2024 14:18	SMPL
P4831-01	E29S4	-80	-0.0535 -		SFAM01.1	PPB	11/26/2024 14:21	SMPL
P4831-02	E29S4D	-48	-0.0456 -		SFAM01.1	PPB	11/26/2024 14:23	SMPL
P4831-03	E29S4S	4354	1.053 -		SFAM01.1	PPB	11/26/2024 14:25	SMPL
P4831-04	E29T2	0	-0.0336 -		SFAM01.1	PPB	11/26/2024 14:27	SMPL
P4831-05	E29T4	2	-0.0331 -		SFAM01.1	PPB	11/26/2024 14:30	SMPL
P4831-06	E29W0	324	0.0473 -		SFAM01.1	PPB	11/26/2024 14:32	SMPL
P4831-07	E29W4	154	0.0048 -		SFAM01.1	PPB	11/26/2024 14:34	SMPL
P4831-08	E29W1	174	0.0098 -		SFAM01.1	PPB	11/26/2024 14:37	SMPL
P4831-09	E29W5	170	0.0088 -		SFAM01.1	PPB	11/26/2024 14:39	SMPL
P4831-10	E29R6	50	-0.0211 -		SFAM01.1	PPB	11/26/2024 14:41	SMPL
P4831-11	E29S0	35	-0.0248 -		SFAM01.1	PPB	11/26/2024 14:43	SMPL
P4831-12	E29S2	65	-0.0174 -		SFAM01.1	PPB	11/26/2024 14:46	SMPL
P4831-13	E29S6	81	-0.0134 -		SFAM01.1	PPB	11/26/2024 14:48	SMPL
P4831-14	E29W6	164	0.0073 -		SFAM01.1	PPB	11/26/2024 14:50	SMPL
P4831-15	E29W2	101	-0.0084 -		SFAM01.1	PPB	11/26/2024 14:52	SMPL
P4832-01	E29S5	143	0.0021 -		SFAM01.1	PPB	11/26/2024 14:55	SMPL
P4832-02	E29S5D	74	-0.0151 -		SFAM01.1	PPB	11/26/2024 14:57	SMPL
P4832-03	E29S5S	5064	1.2302 -		SFAM01.1	PPB	11/26/2024 14:59	SMPL
P4832-04	E29T3	73	-0.0154 -		SFAM01.1	PPB	11/26/2024 15:01	SMPL
CCV056	CCV056	19406	4.8094 -		SFAM01.1	PPB	11/26/2024 15:04	SMPL
CCB056	CCB056	-193	-0.0817 -		SFAM01.1	PPB	11/26/2024 15:06	SMPL
P4832-05	E29T5	252	0.0293 -		SFAM01.1	PPB	11/26/2024 15:08	SMPL
P4832-06	E29R7	161	0.0066 -		SFAM01.1	PPB	11/26/2024 15:10	SMPL
P4832-07	E29S1	407	0.068 -		SFAM01.1	PPB	11/26/2024 15:13	SMPL
P4832-08	E29S3	238	0.0258 -		SFAM01.1	PPB	11/26/2024 15:15	SMPL
P4832-09	E29S7	257	0.0306 -		SFAM01.1	PPB	11/26/2024 15:17	SMPL
PB165276BL	PBW276	108	-0.0066 -		SFAM01.1	PPB	11/26/2024 15:20	SMPL
P4877-01	E29Z1	175	0.0101 -		SFAM01.1	PPB	11/26/2024 15:22	SMPL
P4877-02	E29Z3	41	-0.0234 -		SFAM01.1	PPB	11/26/2024 15:24	SMPL
P4877-03	E29Z5	65	-0.0174 -		SFAM01.1	PPB	11/26/2024 15:26	SMPL
P4877-04	E29Z7	343	0.052 -		SFAM01.1	PPB	11/26/2024 15:29	SMPL
P4877-05	E2A13	247	0.0281 -		SFAM01.1	PPB	11/26/2024 15:31	SMPL
P4877-06	E2A19	311	0.044 -		SFAM01.1	PPB	11/26/2024 15:33	SMPL
P4877-07	E2A22	189	0.0136 -		SFAM01.1	PPB	11/26/2024 15:35	SMPL
P4877-08	E2A23	250	0.0288 -		SFAM01.1	PPB	11/26/2024 15:38	SMPL

LB133633

INSTRUMENT ID : CV1

P4877-09	E29Z9	253	0.0296 -	SFAM01.1	PPB	11/26/2024 15:40 SMPL
P4877-10	E2A01	323	0.047 -	SFAM01.1	PPB	11/26/2024 15:42 SMPL
P4877-11	E2A01D	131	-0.0009 -	SFAM01.1	PPB	11/26/2024 15:44 SMPL
P4877-12	E2A01S	4581	1.1097 -	SFAM01.1	PPB	11/26/2024 15:47 SMPL
P4877-13	E2A03	224	0.0223 -	SFAM01.1	PPB	11/26/2024 15:49 SMPL
P4877-14	E2A05	239	0.0261 -	SFAM01.1	PPB	11/26/2024 15:51 SMPL
CCV057	CCV057	18931	4.6908 -	SFAM01.1	PPB	11/26/2024 15:54 SMPL
CCB057	CCB057	-116	-0.0625 -	SFAM01.1	PPB	11/26/2024 15:56 SMPL
P4877-15	E2A07	203	0.0171 -	SFAM01.1	PPB	11/26/2024 15:58 SMPL
P4877-16	E2A09	180	0.0113 -	SFAM01.1	PPB	11/26/2024 16:00 SMPL
P4877-17	E2A20	159	0.0061 -	SFAM01.1	PPB	11/26/2024 16:03 SMPL
P4877-18	E2A24	127	-0.0019 -	SFAM01.1	PPB	11/26/2024 16:05 SMPL
P4877-19	E29S9	144	0.0024 -	SFAM01.1	PPB	11/26/2024 16:07 SMPL
PB165277BL	PBW277	340	0.0513 -	SFAM01.1	PPB	11/26/2024 16:09 SMPL
P4878-01	E29Z2	366	0.0578 -	SFAM01.1	PPB	11/26/2024 16:12 SMPL
P4878-02	E29Z4	228	0.0233 -	SFAM01.1	PPB	11/26/2024 16:14 SMPL
P4878-03	E29Z6	206	0.0178 -	SFAM01.1	PPB	11/26/2024 16:16 SMPL
P4878-04	E29Z8	173	0.0096 -	SFAM01.1	PPB	11/26/2024 16:18 SMPL
P4878-05	E2A14	179	0.0111 -	SFAM01.1	PPB	11/26/2024 16:21 SMPL
P4878-06	E2A00	186	0.0128 -	SFAM01.1	PPB	11/26/2024 16:23 SMPL
P4878-07	E2A02	241	0.0266 -	SFAM01.1	PPB	11/26/2024 16:25 SMPL
P4878-08	E2A02D	302	0.0418 -	SFAM01.1	PPB	11/26/2024 16:28 SMPL
P4878-09	E2A02S	5155	1.2529 -	SFAM01.1	PPB	11/26/2024 16:30 SMPL
P4878-10	E2A04	69	-0.0164 -	SFAM01.1	PPB	11/26/2024 16:32 SMPL
P4878-11	E2A06	176	0.0103 -	SFAM01.1	PPB	11/26/2024 16:34 SMPL
P4878-12	E2A08	272	0.0343 -	SFAM01.1	PPB	11/26/2024 16:37 SMPL
P4878-13	E2A10	330	0.0488 -	SFAM01.1	PPB	11/26/2024 16:39 SMPL
CCV058	CCV058	19116	4.737 -	SFAM01.1	PPB	11/26/2024 16:41 SMPL
CCB058	CCB058	-114	-0.062 -	SFAM01.1	PPB	11/26/2024 16:43 SMPL