

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: ENTA05
 ProjectID: North Point
 Lab Code: CHEM Case No.: P4839 SAS No.: P4839 SDG No.: P4839

Calibration Sequence : FF111324		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
500	57168580	114337	FF015003.D
200	24056486	120282	FF015004.D
100	13574431	135744	FF015005.D
50	6559493	131190	FF015006.D
1000	129538208	129538	FF015007.D
AVG RF : 126218		% RSD : 6.894	AVG RT : 15.0152