

SOIL DIESEL RANGE ORGANICS SURROGATE RECOVERY

Lab Name: Chemtech Client: ENTACT
Lab Code: CHEM Case No.: P4839 SAS No.: P4839 SDG No.: P4839

EPA SAMPLE NO.	S1 TETRACOSANE-d50	S2	S3	S4	TOT OUT
PIBLK-FF015077.D	91				0
PIBLK-FF015082.D	96				0
PIBLK-FF015089.D	89				0
PIBLK-FG014800.D	89				0
PIBLK-FG014814.D	89				0
PIBLK-FG014817.D	91				0
PIBLK-FG014832.D	94				0
PIBLK-FG014836.D	95				0
PIBLK-FG014839.D	98				0
PIBLK-FG014847.D	99				0
PIBLK-FG014858.D	88				0
EX-9-TPH-9	59				0
EX-9-TPH-10	68				0
EX-9-TPH-11	61				0
EX-9-TPH-12	60				0
EX-9-TPH-13	68				0
EX-9-TPH-14	0 *				1
EX-9-TPH-15	53				0
EX-9-TPH-16	59				0
EX-9-TPH-17	66				0
EX-9-TPH-18	76				0
EX-9-TPH-18MS	51				0
EX-9-TPH-18MSD	51				0
EX-9-TPH-19	98				0
EX-9-TPH-20	69				0
EX-9-TPH-21	67				0
EX-10-TPH-1	51				0
EX-10-TPH-2	75				0
EX-10-TPH-3	78				0
EX-4-TPH-1	62				0
EX-4-TPH-2	63				0
EX-4-TPH-3	60				0
EX-4-TPH-4	51				0
EX-4-TPH-5	0 *				1
EX-4-TPH-6	81				0
EX-4-TPH-6MS	80				0
EX-4-TPH-6MSD	82				0

QC LIMITS

TETRACOSANE-d50

For Water : 29-130
For Soil : 37-130

Column to be used to flag recovery values
* Values outside of contract required QC limits
D Surrogate Diluted Out

SOIL DIESEL RANGE ORGANICS SURROGATE RECOVERY

Lab Name: Chemtech Client: ENTACT
Lab Code: CHEM Case No.: P4839 SAS No.: P4839 SDG No.: P4839

EPA SAMPLE NO.	S1 TETRACOSANE-d50	S2	S3	S4	TOT OUT
PB164995BL	84				0
PB164995BS	89				0
PB165028BL	83				0
PB165028BS	90				0

QC LIMITS

TETRACOSANE-d50

For Water : 29-130
For Soil : 37-130

Column to be used to flag recovery values
* Values outside of contract required QC limits
D Surrogate Diluted Out