

SOIL DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RI

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P4839 **SAS No :** P4839 **SDG No:** P4839
Matrix Spike - EPA Sample No : PB165028BS **Datafile:** FF015085.D

COMPOUND	SPIKE ADDED ug/kg	CONCENTRATION ug/kg	LCS/LCSD CONCENTRATION ug/kg	% REC	QC LIMITS
DRO	6662	0	8340	125	68-131