

Continuing Calibration Report for SequenceID : FC111524AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067770.D

Aliphatic C9-C12	9927990.000	60.000	3.172	6.448	165466.500	165791.359	0.196
Aliphatic C12-C16	6889494.000	40.000	6.449	9.838	172237.350	170384.422	-1.087
Aliphatic C16-C21	10337141.000	60.000	9.839	13.196	172285.683	169621.553	-1.571
Aliphatic C21-C28	12369366.000	80.000	13.197	16.850	154617.075	157280.060	1.693
Aliphatic C28-C40	13784773.000	120.000	16.851	21.693	114873.108	121111.336	5.151
Aliphatic EPH	53308764.000	360.000	3.172	21.693	148079.900	150155.324	1.382