

Continuing Calibration Report for SequenceID : FC111524AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067785.D

Aliphatic C9-C12	10061913.000	60.000	3.172	6.448	167698.550	165791.359	-1.150
Aliphatic C12-C16	6933044.000	40.000	6.449	9.838	173326.100	170384.422	-1.726
Aliphatic C16-C21	10262529.000	60.000	9.839	13.196	171042.150	169621.553	-0.838
Aliphatic C21-C28	12272715.000	80.000	13.197	16.850	153408.938	157280.060	2.461
Aliphatic C28-C40	13658188.000	120.000	16.851	21.693	113818.233	121111.336	6.022
Aliphatic EPH	53188389.000	360.000	3.172	21.693	147745.525	150155.324	1.605