

Continuing Calibration Report for SequenceID : FC111824AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067823.D

Aliphatic C9-C12	9576563.000	60.000	3.172	6.448	159609.383	165791.359	3.729
Aliphatic C12-C16	6646291.000	40.000	6.449	9.838	166157.275	170384.422	2.481
Aliphatic C16-C21	9638248.000	60.000	9.839	13.196	160637.467	169621.553	5.297
Aliphatic C21-C28	12169776.000	80.000	13.197	16.850	152122.200	157280.060	3.279
Aliphatic C28-C40	13421842.000	120.000	16.851	21.691	111848.683	121111.336	7.648
Aliphatic EPH	51452720.000	360.000	3.172	21.691	142924.222	150155.324	4.816