

Continuing Calibration Report for SequenceID : FD111524AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048754.D

Aromatic C10-C12	7741682.000	40.000	4.088	5.805	193542.050	192017.544	-0.794
Aromatic C12-C16	11683206.000	60.000	5.806	8.411	194720.100	192458.977	-1.175
Aromatic C16-C21	14938916.000	80.000	8.412	12.674	186736.450	185804.816	-0.501
Aromatic C21-C36	28774591.000	180.000	12.675	18.081	159858.839	157231.056	-1.671
Aromatic EPH	63138395.000	360.000	4.088	18.081	175384.431	173317.266	-1.193