

Continuing Calibration Report for SequenceID : FD111824AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048765.D

Aromatic C10-C12	7925969.000	40.000	4.088	5.805	198149.225	192017.544	-3.193
Aromatic C12-C16	11827424.000	60.000	5.806	8.411	197123.733	192458.977	-2.424
Aromatic C16-C21	14986986.000	80.000	8.412	12.674	187337.325	185804.816	-0.825
Aromatic C21-C36	28594859.000	180.000	12.675	18.081	158860.328	157231.056	-1.036
Aromatic EPH	63335238.000	360.000	4.088	18.081	175931.217	173317.266	-1.508