

Continuing Calibration Report for SequenceID : FD111824AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048780.D

Aromatic C10-C12	8154443.000	40.000	4.088	5.805	203861.075	192017.544	-6.168
Aromatic C12-C16	12059712.000	60.000	5.806	8.411	200995.200	192458.977	-4.435
Aromatic C16-C21	15225694.000	80.000	8.412	12.674	190321.175	185804.816	-2.431
Aromatic C21-C36	29380056.000	180.000	12.675	18.081	163222.533	157231.056	-3.811
Aromatic EPH	64819905.000	360.000	4.088	18.081	180055.292	173317.266	-3.888