

Continuing Calibration Report for SequenceID : FD111824AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048783.D

Aromatic C10-C12	8433654.000	40.000	4.088	5.805	210841.350	192017.544	-9.803
Aromatic C12-C16	12609540.000	60.000	5.806	8.411	210159.000	192458.977	-9.197
Aromatic C16-C21	15976608.000	80.000	8.412	12.674	199707.600	185804.816	-7.482
Aromatic C21-C36	30236202.000	180.000	12.675	18.081	167978.900	157231.056	-6.836
Aromatic EPH	67256004.000	360.000	4.088	18.081	186822.233	173317.266	-7.792