

Continuing Calibration Report for SequenceID : FE111524AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE051261.D

Aliphatic C9-C12	10149366.000	60.000	3.134	6.753	169156.100	139914.732	-20.899
Aliphatic C12-C16	6608362.000	40.000	6.754	10.185	165209.050	140635.568	-17.473
Aliphatic C16-C21	9525547.000	60.000	10.186	13.544	158759.117	137768.027	-15.237
Aliphatic C21-C28	12166198.000	80.000	13.545	17.200	152077.475	133951.574	-13.532
Aliphatic C28-C40	17018018.000	120.000	17.201	22.043	141816.817	128748.418	-10.150
Aliphatic EPH	55467491.000	360.000	3.134	22.043	154076.364	134589.790	-14.478