

Continuing Calibration Report for SequenceID : FE111524AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE051276.D

Aliphatic C9-C12	9208781.000	60.000	3.134	6.753	153479.683	139914.732	-9.695
Aliphatic C12-C16	6047179.000	40.000	6.754	10.185	151179.475	140635.568	-7.497
Aliphatic C16-C21	8792528.000	60.000	10.186	13.544	146542.133	137768.027	-6.369
Aliphatic C21-C28	11435247.000	80.000	13.545	17.200	142940.588	133951.574	-6.711
Aliphatic C28-C40	15501398.000	120.000	17.201	22.043	129178.317	128748.418	-0.334
Aliphatic EPH	50985133.000	360.000	3.134	22.043	141625.369	134589.790	-5.227