

Continuing Calibration Report for SequenceID : FE111924AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE051337.D

Aliphatic C28-C40	15751632.000	120.000	17.199	22.041	131263.600	128748.418	-1.954
Aliphatic EPH	53278076.000	360.000	3.136	22.041	147994.656	134589.790	-9.960
Aliphatic C9-C12	10147030.000	60.000	3.136	6.752	169117.167	139914.732	-20.872
Aliphatic C12-C16	6396528.000	40.000	6.753	10.184	159913.200	140635.568	-13.708
Aliphatic C16-C21	8947952.000	60.000	10.185	13.543	149132.533	137768.027	-8.249
Aliphatic C21-C28	12034934.000	80.000	13.544	17.198	150436.675	133951.574	-12.307