

Continuing Calibration Report for SequenceID : FE111924AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE051345.D

Aliphatic C28-C40	15991458.000	120.000	17.199	22.041	133262.150	128748.418	-3.506
Aliphatic EPH	51471892.000	360.000	3.136	22.041	142977.478	134589.790	-6.232
Aliphatic C9-C12	9574535.000	60.000	3.136	6.752	159575.583	139914.732	-14.052
Aliphatic C12-C16	6063357.000	40.000	6.753	10.184	151583.925	140635.568	-7.785
Aliphatic C16-C21	8548935.000	60.000	10.185	13.543	142482.250	137768.027	-3.422
Aliphatic C21-C28	11293607.000	80.000	13.544	17.198	141170.088	133951.574	-5.389