

Continuing Calibration Report for SequenceID : FF111624AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015068.D

Aromatic C10-C12	5493316.000	40.000	4.507	6.371	137332.900	136785.903	-0.400
Aromatic C12-C16	8072512.000	60.000	6.372	9.062	134541.867	135022.207	0.356
Aromatic C16-C21	10039777.000	80.000	9.063	13.374	125497.213	127893.613	1.874
Aromatic C21-C36	21625976.000	180.000	13.375	18.817	120144.311	121872.475	1.418
Aromatic EPH	45231581.000	360.000	4.507	18.817	125643.281	127059.175	1.114