

Continuing Calibration Report for SequenceID : FF111624AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015075.D

Aromatic C10-C12	5613857.000	40.000	4.507	6.371	140346.425	136785.903	-2.603
Aromatic C12-C16	8279747.000	60.000	6.372	9.062	137995.783	135022.207	-2.202
Aromatic C16-C21	10353508.000	80.000	9.063	13.374	129418.850	127893.613	-1.193
Aromatic C21-C36	22031721.000	180.000	13.375	18.817	122398.450	121872.475	-0.432
Aromatic EPH	46278833.000	360.000	4.507	18.817	128552.314	127059.175	-1.175