

SOLID EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P4839 **SAS No :** P4839 **SDG No:** P4839
Sample No : P4839-07MS **Datafile:** FC067776.D
Client ID : EX-9-TPH-15MS

COMPOUND	SPIKE ADDED mg/kg	SAMPLE CONCENTRATION mg/kg	MS/MSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	12.1	26.3	45.6	159	*	(40-140)
Aliphatic C12-C16	8.1	38.7	67.6	358	*	(40-140)
Aliphatic C16-C21	12.1	2.39	16.9	120		(40-140)
Aliphatic C21-C28	16.1	0	18.0	111		(40-140)
Aliphatic C28-C40	24.2	3.38	33.4	124		(40-140)