

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P4839 **SAS No :** P4839 **SDG No:** P4839
Sample No : PB164996BSD **Datafile:** FC067773.D
Client ID : PB164996BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	10	5.33	53		10 (40-140)	50
Aliphatic C12-C16	6.7	5.16	78		8.6 (40-140)	50
Aliphatic C16-C21	10	9.20	92		15 (40-140)	50
Aliphatic C21-C28	13.3	12.3	92		14 (40-140)	50
Aliphatic C28-C40	20.0	21.7	108		18 (40-140)	50