

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P4839 **SAS No :** P4839 **SDG No:** P4839
Sample No : PB165032BS **Datafile:** FC067808.D
Client ID : PB165032BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	10	6.39	64		(40-140)
Aliphatic C12-C16	6.7	5.31	80		(40-140)
Aliphatic C16-C21	10	8.07	81		(40-140)
Aliphatic C21-C28	13.3	10.4	78		(40-140)
Aliphatic C28-C40	20.0	17.7	89		(40-140)