

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P4839 **SAS No :** P4839 **SDG No:** P4839
Sample No : PB165032BS **Datafile:** FD048767.D
Client ID : PB165032BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aromatic C10-C12	6.7	5.16	78		(40-140)
Aromatic C12-C16	10	9.14	92		(40-140)
Aromatic C16-C21	16.7	16.7	100		(40-140)
Aromatic C21-C36	26.6	26.3	99		(40-140)