

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4843

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133442

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10	10	100	90-110	10/16/2024
Chloride	mg/L	3	3	100	90-110	10/16/2024
Fluoride	mg/L	2	2	100	90-110	10/16/2024
Nitrite	mg/L	3	3	100	90-110	10/16/2024
Nitrate	mg/L	2.5	2.5	100	90-110	10/16/2024
Sulfate	mg/L	14.9	15	99	90-110	10/16/2024
Orthophosphate as P	mg/L	4.8	5	96	90-110	10/16/2024
Sample ID: CCV1						
Bromide	mg/L	10.4	10	104	90-110	11/14/2024
Chloride	mg/L	3.1	3	103	90-110	11/14/2024
Fluoride	mg/L	2.1	2	105	90-110	11/14/2024
Nitrite	mg/L	3.1	3	103	90-110	11/14/2024
Nitrate	mg/L	2.6	2.5	104	90-110	11/14/2024
Sulfate	mg/L	15.5	15	103	90-110	11/14/2024
Orthophosphate as P	mg/L	5.3	5	106	90-110	11/14/2024
Sample ID: CCV2						
Bromide	mg/L	9.7	10	97	90-110	11/14/2024
Chloride	mg/L	2.9	3	97	90-110	11/14/2024
Fluoride	mg/L	2	2	100	90-110	11/14/2024
Nitrite	mg/L	3	3	100	90-110	11/14/2024
Nitrate	mg/L	2.4	2.5	96	90-110	11/14/2024
Sulfate	mg/L	14.6	15	97	90-110	11/14/2024
Orthophosphate as P	mg/L	4.9	5	98	90-110	11/14/2024

Initial and Continuing Calibration Verification**Client:** ENTACT**SDG No.:** P4843**Project:** 540 Degraw St, Brooklyn, NY - E9309**RunNo.:** LB133444

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.00	7	100	90-110	11/14/2024
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	11/14/2024
Sample ID:	CCV2						
pH		pH	12.01	12.00	100	90-110	11/14/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4843

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133448

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.504	0.5	101	90-110	11/14/2024
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	11/14/2024
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.503	0.5	101	90-110	11/14/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4843

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133474

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV						
Flash Point	o F	82.1	81	101	78-84	11/15/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P4843

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133557

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 TKN	mg/L	4.6	5	92	90-110	11/21/2024
Sample ID: CCV1 TKN	mg/L	5	5	100	90-110	11/21/2024
Sample ID: CCV2 TKN	mg/L	5	5	100	90-110	11/21/2024
Sample ID: CCV3 TKN	mg/L	5	5	100	90-110	11/21/2024