

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	P4843
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	P4834-01
Client ID:	SLP-1-WATERDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	7.09		7.10		1	0.14		11/14/2024

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	P4843
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	P4840-02
Client ID:	COMPDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	728		734		1	0.82		11/15/2024

Duplicate Sample Summary

Client: ENTACT Project: 540 Degraw St, Brooklyn, NY - E9309 Client ID: SW-WTS-01DUP	SDG No.: P4843 Sample ID: P4843-01 Percent Solids for Spike Sample: 0
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TS	mg/L	+/-5	1710		1750		1	2.43		11/14/2024

Duplicate Sample Summary

Client: ENTACT	SDG No.: P4843
Project: 540 Degraw St, Brooklyn, NY - E9309	Sample ID: P4843-01
Client ID: SW-WTS-01MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Fluoride	mg/L	+/-20	2.50		2.50		1	0		11/14/2024
Nitrate	mg/L	+/-20	3.60		3.60		1	0		11/14/2024
Nitrite	mg/L	+/-20	3.20		3.20		1	0		11/14/2024
Chloride	mg/L	+/-20	26.8	OR	26.8	OR	1	0		11/14/2024
Sulfate	mg/L	+/-20	532	OR	531	OR	1	0		11/14/2024
Bromide	mg/L	+/-20	11.4		11.3		1	1		11/14/2024
Orthophosphate as P	mg/L	+/-20	0.80	J	1.20		1	40	*	11/14/2024

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	P4843
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	P4853-02
Client ID:	002-35TH-AVE(NOV)DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	44.2		45.6		1	3.08		11/14/2024

Duplicate Sample Summary

Client:	ENTACT	SDG No.:	P4843
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	P4871-01
Client ID:	WC-10-A-202411DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Flash Point	o F	+/-2	>212.0		>212.0		1	0		11/15/2024