

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: P4843

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB133442

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133442BSW							
Bromide	mg/L	10	10.3		103	1	90-110	11/14/2024
Chloride	mg/L	3	3.10		103	1	90-110	11/14/2024
Fluoride	mg/L	2	2.10		105	1	90-110	11/14/2024
Nitrite	mg/L	3	3.10		103	1	90-110	11/14/2024
Nitrate	mg/L	2.5	2.60		104	1	90-110	11/14/2024
Sulfate	mg/L	15	15.2		101	1	90-110	11/14/2024
Orthophosphate as P	mg/L	5	5.10		102	1	90-110	11/14/2024

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: P4843

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB133448

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	lb133448BS							
Hexavalent Chromium	mg/L	0.5	0.51		103	1	90-111	11/14/2024

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: P4843

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB133454

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133454BS							
BOD5	mg/L	198	184		93	1	84.6-115.4	11/14/2024

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: P4843

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB133467

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133467BS							
TSS	mg/L	550	544		99	1	90-110	11/15/2024

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: P4843

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB133557

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165161BS							
TKN	mg/L	5	4.90		98	1	90-110	11/21/2024