



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** P4843
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4843 **SAS No.:** P4843
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	93	40 - 160	CV	11/14/2024	12:50	LB133446
CRI01	Cadmium	6.88	6.0	115	40 - 160	P	11/15/2024	12:05	LB133471
	Copper	22.1	20.0	110	40 - 160	P	11/15/2024	12:05	LB133471
	Lead	12.2	12.0	101	40 - 160	P	11/15/2024	12:05	LB133471
	Nickel	38.9	40.0	97	40 - 160	P	11/15/2024	12:05	LB133471
	Zinc	44.2	40.0	110	40 - 160	P	11/15/2024	12:05	LB133471