

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: ENTACT **SDG No.:** P4843
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4843 **SAS No.:** P4843
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	6.99	1.0	699	-5	7	11/15/2024	12:09	LB133471
	Copper	5.35	2.0	268	-18	22	11/15/2024	12:09	LB133471
	Lead	8.65			-12	12	11/15/2024	12:09	LB133471
	Nickel	1.62	2.0	81	-38	42	11/15/2024	12:09	LB133471
	Zinc	8.16			-40	40	11/15/2024	12:09	LB133471
ICSAB01	Cadmium	1000	972	103	826	1120	11/15/2024	12:17	LB133471
	Copper	496	511	97	434	588	11/15/2024	12:17	LB133471
	Lead	55.6	49.0	114	37	61	11/15/2024	12:17	LB133471
	Nickel	1000	954	105	810	1100	11/15/2024	12:17	LB133471
	Zinc	894	952	94	809	1095	11/15/2024	12:17	LB133471