

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** P4843
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4843 **SAS No.:** P4843

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164972BS							
Cadmium	ug/L	100	102		102	80 - 120	P
Copper	ug/L	150	166		111	80 - 120	P
Lead	ug/L	500	524		105	80 - 120	P
Nickel	ug/L	250	262		105	80 - 120	P
Zinc	ug/L	100	114		114	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: P4843

Contract: ENTA05

Lab Code: CHEM

Case No.: P4843

SAS No.: P4843

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164975BS Mercury	ug/L	4.0	4.22		106	80 - 120	CV