

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4843 SAS No.: P4843 SDG NO.: P4843

Continuing Calib Date: 11/16/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 01:34 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.78	6.78	6.68	6.88	0.00
Aroclor-1254-2	(2)	7.00	7.00	6.90	7.10	0.01
Aroclor-1254-3	(3)	7.36	7.36	7.26	7.46	0.00
Aroclor-1254-4	(4)	7.64	7.64	7.54	7.74	0.00
Aroclor-1254-5	(5)	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene		4.76	4.76	4.66	4.86	0.00
Decachlorobiphenyl		10.68	10.67	10.57	10.77	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4843 SAS No.: P4843 SDG NO.: P4843

Continuing Calib Date: 11/16/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 01:34 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.97	5.97	5.87	6.07	0.00
Aroclor-1254-2	(2)	6.12	6.12	6.02	6.22	0.00
Aroclor-1254-3	(3)	6.53	6.53	6.43	6.63	0.00
Aroclor-1254-4	(4)	6.75	6.76	6.66	6.86	0.01
Aroclor-1254-5	(5)	7.18	7.18	7.08	7.28	0.00
Tetrachloro-m-xylene		4.05	4.05	3.95	4.15	0.00
Decachlorobiphenyl		9.21	9.21	9.11	9.31	0.00