

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4843 SAS No.: P4843 SDG NO.: P4843

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL02 Date Analyzed: 11/16/2024

Lab Sample No.: AR1660CCC500 Data File : PP068524.D Time Analyzed: 00:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.924	5.821	6.021	467.470	500.000	-6.5
Aroclor-1016-2	5.946	5.844	6.044	474.360	500.000	-5.1
Aroclor-1016-3	6.010	5.908	6.108	481.730	500.000	-3.7
Aroclor-1016-4	6.108	6.005	6.205	487.190	500.000	-2.6
Aroclor-1016-5	6.402	6.299	6.499	471.960	500.000	-5.6
Aroclor-1260-1	7.526	7.423	7.623	449.420	500.000	-10.1
Aroclor-1260-2	7.779	7.676	7.876	443.760	500.000	-11.2
Aroclor-1260-3	8.141	8.038	8.238	440.350	500.000	-11.9
Aroclor-1260-4	8.379	8.276	8.476	430.390	500.000	-13.9
Aroclor-1260-5	8.717	8.613	8.813	426.140	500.000	-14.8
Decachlorobiphenyl	10.676	10.572	10.772	44.620	50.000	-10.8
Tetrachloro-m-xylene	4.760	4.657	4.857	49.160	50.000	-1.7

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P4843 SAS No.: P4843 SDG NO.: P4843

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL02 Date Analyzed: 11/16/2024

Lab Sample No.: AR1660CCC500 Data File : PP068524.D Time Analyzed: 00:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.157	5.058	5.258	503.140	500.000	0.6
Aroclor-1016-2	5.177	5.079	5.279	488.860	500.000	-2.2
Aroclor-1016-3	5.357	5.259	5.459	503.750	500.000	0.8
Aroclor-1016-4	5.397	5.298	5.498	504.320	500.000	0.9
Aroclor-1016-5	5.616	5.517	5.717	479.000	500.000	-4.2
Aroclor-1260-1	6.659	6.560	6.760	463.200	500.000	-7.4
Aroclor-1260-2	6.846	6.747	6.947	450.770	500.000	-9.8
Aroclor-1260-3	7.003	6.905	7.105	458.030	500.000	-8.4
Aroclor-1260-4	7.477	7.378	7.578	463.390	500.000	-7.3
Aroclor-1260-5	7.717	7.617	7.817	444.670	500.000	-11.1
Decachlorobiphenyl	9.209	9.110	9.310	45.170	50.000	-9.7
Tetrachloro-m-xylene	4.051	3.951	4.151	49.890	50.000	-0.2