

Surrogate Summary

SDG No.: **P4843**

Client: **ENTACT**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP068360.D	PIBLK-PP068360.D	Tetrachloro-m-xylene	1	20	21.0	105		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	23.4	117		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	21.6	108		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	23.6	118		70 (60)	130 (140)
I.BLK-PP068514.D	PIBLK-PP068514.D	Tetrachloro-m-xylene	1	20	20.5	103		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	21.0	105		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	21.4	107		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	20.3	101		70 (60)	130 (140)
PB164990BL	PB164990BL	Tetrachloro-m-xylene	1	20	19.4	97		30 (10)	150 (157)
		Decachlorobiphenyl	1	20	19.9	100		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	20.1	100		30 (10)	150 (157)
		Decachlorobiphenyl	2	20	20.1	100		30 (10)	150 (173)
PB164990BS	PB164990BS	Tetrachloro-m-xylene	1	20	20.1	101		30 (10)	150 (157)
		Decachlorobiphenyl	1	20	20.4	102		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	20.7	104		30 (10)	150 (157)
		Decachlorobiphenyl	2	20	20.1	100		30 (10)	150 (173)
PB164990BSD	PB164990BSD	Tetrachloro-m-xylene	1	20	19.8	99		30 (10)	150 (157)
		Decachlorobiphenyl	1	20	20.1	101		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	20.2	101		30 (10)	150 (157)
		Decachlorobiphenyl	2	20	20.1	100		30 (10)	150 (173)
P4843-01	SW-WTS-01	Tetrachloro-m-xylene	1	20	13.4	67		30 (10)	150 (157)
		Decachlorobiphenyl	1	20	6.99	35		30 (10)	150 (173)
		Tetrachloro-m-xylene	2	20	13.9	69		30 (10)	150 (157)
		Decachlorobiphenyl	2	20	6.83	34		30 (10)	150 (173)
I.BLK-PP068528.D	PIBLK-PP068528.D	Tetrachloro-m-xylene	1	20	20.3	102		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	21.5	108		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	21.6	108		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	22.2	111		70 (60)	130 (140)