

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4843

Client: ENTACT

Analytical Method: 8270E DataFile: BF140490.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB165039BS	Phenol	50	43.2	ug/L	86				20 (66)	160 (118)	
	1,4-Dichlorobenzene	50	42.4	ug/L	85				70 (76)	130 (103)	
	1,2,4-Trichlorobenzene	50	41.5	ug/L	83				70 (41)	130 (115)	
	Naphthalene	50	42.9	ug/L	86				70 (64)	130 (107)	