

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4843

Client: ENTACT

Analytical Method: 8270E DataFile: BF140491.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD Qual	Low	Limits	
										High	RPD
PB165039BSD	Phenol	50	42.4	ug/L	85	2			20 (66)	160 (118)	20 (20)
	1,4-Dichlorobenzene	50	41.8	ug/L	84	1			70 (76)	130 (103)	20 (20)
	1,2,4-Trichlorobenzene	50	41.3	ug/L	83	0			70 (41)	130 (115)	20 (20)
	Naphthalene	50	43.0	ug/L	86	0			70 (64)	130 (107)	20 (20)