

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103124\
 Data File : VN084615.D
 Acq On : 31 Oct 2024 12:56
 Operator : JC\MD
 Sample : VSTDICC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC050

Quant Time: Nov 01 02:22:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N103124W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 01 02:19:19 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/04/2024
 Supervised By : Mahesh Dadoda 11/04/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	32918	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	176826	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	164720	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	78237	29.563	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery =	98.533%		
60) 4-Bromofluorobenzene	12.847	95	87434	30.828	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery =	102.767%		
63) Toluene-d8	10.565	98	233304	29.919	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery =	99.733%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	94735	49.481	ug/l	96
3) Chloromethane	2.359	50	109029	47.299	ug/l	98
4) Vinyl Chloride	2.512	62	106283	49.654	ug/l	97
5) Bromomethane	2.901	94	51725	47.610	ug/l	100
6) Chloroethane	3.077	64	71616m	50.344	ug/l	
7) Trichlorofluoromethane	3.471	101	169605	48.790	ug/l	97
8) Diethyl Ether	3.953	74	61891	49.913	ug/l	96
9) 1,1,2-Trichlorotrifluoro...	4.359	101	97991	49.880	ug/l	99
10) 1,1-Dichloroethene	4.324	96	93892	48.943	ug/l	95
11) Methyl Iodide	4.583	142	130322	49.426	ug/l	100
12) Methyl Acetate	5.018	43	127594	42.624	ug/l	94
13) Acrolein	4.171	56	85754	232.550	ug/l	98
14) Acrylonitrile	5.712	53	243950	249.373	ug/l	99
15) Acetone	4.430	58	61279	243.784	ug/l	91
16) Carbon Disulfide	4.700	76	275193	48.752	ug/l	99
17) Allyl chloride	5.012	41	152844	47.220	ug/l	98
18) Methylene Chloride	5.265	84	108652	48.332	ug/l	98
19) trans-1,2-Dichloroethene	5.783	96	99291	49.378	ug/l	91
20) Diisopropyl ether	6.671	45	332986	50.885	ug/l	98
21) 1,1-Dichloroethane	6.565	63	190325	48.363	ug/l	98
22) cis-1,2-Dichloroethene	7.483	96	119146	49.630	ug/l	94
23) tert-Butyl Alcohol	5.542	59	75243	220.317	ug/l #	100
24) Methyl tert-Butyl Ether	5.795	73	308206	49.680	ug/l	99
25) Chloroform	7.959	83	197402	48.959	ug/l	94
26) Cyclohexane	8.253	56	153309	51.000	ug/l #	100
29) 1,1-Dichloropropene	8.371	75	133196	50.010	ug/l	100
30) 2-Butanone	7.477	43	305143	248.373	ug/l	99
31) 2,2-Dichloropropane	7.483	77	161607	49.708	ug/l	98
32) 1,1,1-Trichloroethane	8.165	97	178128	49.229	ug/l	99
33) Carbon Tetrachloride	8.359	117	152324	49.334	ug/l	98
34) Benzene	8.600	78	434965	49.409	ug/l	97
35) Methacrylonitrile	7.777	41	69021	47.401	ug/l	97
36) 1,2-Dichloroethane	8.665	62	143921	49.161	ug/l	100
37) Trichloroethene	9.353	130	97539	49.114	ug/l	99
38) Methylcyclohexane	9.600	83	145824	50.859	ug/l	98
39) 1,2-Dichloropropane	9.618	63	103300	48.746	ug/l	97
40) Dibromomethane	9.706	93	72848	49.139	ug/l	99
41) Bromodichloromethane	9.882	83	154994	48.769	ug/l	99
42) Vinyl Acetate	6.600	43	1161725	252.978	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN103124\
 Data File : VN084615.D
 Acq On : 31 Oct 2024 12:56
 Operator : JC\MD
 Sample : VSTDICC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC050

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 11/04/2024
 Supervised By : Mahesh Dadoda 11/04/2024

Quant Time: Nov 01 02:22:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N103124W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 01 02:19:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.559	43	131344	50.265	ug/l	98
44) Isopropyl Acetate	8.688	43	231400	44.267	ug/l	98
45) 1,4-Dioxane	9.694	88	33980	973.675	ug/l #	89
46) Methyl methacrylate	9.677	41	104945	49.957	ug/l	96
47) n-amyl Acetate	12.494	43	202564	51.886	ug/l	98
48) t-1,3-Dichloropropene	10.835	75	157614	50.224	ug/l	95
49) cis-1,3-Dichloropropene	10.312	75	167077	49.675	ug/l	97
50) 1,1,2-Trichloroethane	11.012	97	98990	49.780	ug/l	96
51) Ethyl methacrylate	10.871	69	158738	51.786	ug/l	96
52) 1,3-Dichloropropane	11.159	76	172748	50.047	ug/l	99
53) Dibromochloromethane	11.359	129	118100	50.750	ug/l	97
54) 1,2-Dibromoethane	11.465	107	99236	49.341	ug/l	96
55) 2-Chloroethyl vinyl ether	10.159	63	395250	263.010	ug/l	99
56) Bromoform	12.576	173	79210	50.781	ug/l	99
58) 4-Methyl-2-Pentanone	10.447	43	647306	249.093	ug/l	98
59) 2-Hexanone	11.194	43	478961	251.701	ug/l	99
61) Tetrachloroethene	11.100	164	85090	49.359	ug/l	98
62) Toluene	10.629	91	459711	49.000	ug/l	99
64) Chlorobenzene	11.888	112	279544	49.357	ug/l	97
65) 1,1,1,2-Tetrachloroethane	11.959	131	104199	50.060	ug/l	98
66) Ethyl Benzene	11.965	91	502126	50.927	ug/l	94
67) m/p-Xylenes	12.071	106	387469	102.646	ug/l	98
68) o-Xylene	12.394	106	187021	51.361	ug/l	98
69) Styrene	12.412	104	323595	53.002	ug/l	98
70) Isopropylbenzene	12.694	105	470823	52.050	ug/l	99
71) 1,1,2,2-Tetrachloroethane	12.935	83	145223	49.532	ug/l	97
72) 1,2,3-Trichloropropane	12.994	75	117853m	42.135	ug/l	
73) Bromobenzene	12.976	156	116686	50.015	ug/l	97
74) n-propylbenzene	13.035	91	557232	52.681	ug/l	99
75) 2-Chlorotoluene	13.123	91	356739	51.790	ug/l	98
76) 1,3,5-Trimethylbenzene	13.171	105	406282	52.568	ug/l	99
77) t-1,4-Dichloro-2-butene	12.735	75	53054	51.671	ug/l	86
78) 4-Chlorotoluene	13.218	91	357227	51.329	ug/l	100
79) tert-butylbenzene	13.435	119	331066	51.799	ug/l	99
80) 1,2,4-Trimethylbenzene	13.482	105	415493	53.289	ug/l	98
81) sec-Butylbenzene	13.618	105	464362	52.727	ug/l	100
82) p-Isopropyltoluene	13.729	119	396709	53.609	ug/l	99
83) 1,3-Dichlorobenzene	13.729	146	217624	50.438	ug/l	99
84) 1,4-Dichlorobenzene	13.812	146	216965	50.890	ug/l	97
85) n-Butylbenzene	14.053	91	325553	52.381	ug/l	99
86) Hexachloroethane	14.329	117	75612	49.961	ug/l	99
87) 1,2-Dichlorobenzene	14.106	146	215946	50.790	ug/l	98
88) 1,2-Dibromo-3-Chloropr...	14.717	75	28965	51.201	ug/l	96
89) 1,2,4-Trichlorobenzene	15.835	180	104664	50.558	ug/l	96
90) Hexachlorobutadiene	15.494	225	44832	49.577	ug/l	95
91) Naphthalene	15.641	128	349297	52.834	ug/l	99
92) 1,2,3-Trichlorobenzene	15.835	180	104664	50.558	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

