



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111524\
Data File : VN084866.D
Acq On : 15 Nov 2024 13:01
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N103124W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Fri Nov 01 02:38:04 2024



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1844

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.5	4798	PASS
75	95	30	60	53.3	14620	PASS
95	95	100	100	100.0	27421	PASS
96	95	5	9	6.8	1877	PASS
173	174	0.00	2	1.0	208	PASS
174	95	50	100	77.9	21371	PASS
175	174	5	9	6.8	1446	PASS
176	174	95	101	96.0	20520	PASS
177	176	5	9	6.7	1368	PASS