

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GARD04  
 Lab Code: CHEM Case No.: P4844 SAS No.: P4844 SDG NO.: P4844  
 Lab File ID: VN084867.D Date Analyzed: 11/15/2024  
 Instrument ID: MSVOA\_N Time Analyzed: 13:49  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                         | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT # | IS3<br>AREA # | RT #   |
|-------------------------|---------------|-------|---------------|------|---------------|--------|
| 12 HOUR STD             | 29205         | 7.81  | 143711        | 9.10 | 129619        | 11.87  |
| UPPER LIMIT             | 58410         | 8.312 | 287422        | 9.6  | 259238        | 12.365 |
| LOWER LIMIT             | 14602.5       | 7.312 | 71855.5       | 8.6  | 64809.5       | 11.365 |
| EPA SAMPLE NO.          |               |       |               |      |               |        |
| 241113075-01-VOA        | 24378         | 7.81  | 131256        | 9.09 | 127289        | 11.87  |
| 241113050-06-TRIP-BLANK | 28914         | 7.81  | 145852        | 9.10 | 125740        | 11.87  |
| VN1115WBL01             | 20903         | 7.81  | 102633        | 9.10 | 89042         | 11.87  |
| VN1115WBS01             | 28727         | 7.81  | 152762        | 9.10 | 143052        | 11.87  |
| VN1115WBSD01            | 27777         | 7.81  | 143562        | 9.10 | 132133        | 11.87  |

IS1 = Bromochloromethane  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.