

**ALLIANCE** 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

**GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: P4846

MATRIX: Water

METHOD: 9056A,9060A,SM2320 B

- |   | NA | NO | YES |
|---|----|----|-----|
| 1. Blank Contamination - If yes, list compounds and concentrations in each blank:   |    | ✓  |     |
| 2. Matrix Spike Duplicate Recoveries Met Criteria   |    | ✓  |     |
| If not met, list those compounds and their recoveries which fall outside the acceptable range.  |    |    |     |
| The Blank Spike met requirements for all samples. The Matrix Spike (FSND-MW-EVAL-04D-20241113MS) analysis met criteria for all samples except for Chloride due to sample matrix interference. The Matrix Spike (FSND-MW-30-20241114MS) analysis met criteria for all samples except for Chloride due to sample matrix interference. The Matrix Spike Duplicate (FSND-MW-EVAL-04D-20241113MSD) analysis met criteria for all samples except for Chloride due to sample matrix interference. The Matrix Spike Duplicate (FSND-MW-30-20241114MSD) analysis met criteria for all samples except for Chloride due to sample matrix interference. |    |    |     |
| 3. Sample Duplicate Analysis Met QC Criteria  |    |    | ✓   |
| If not met, list those compounds and their recoveries which fall outside the acceptable range.  |    |    |     |
| 4. Digestion Holding Time Met   |    |    | ✓   |
| If not met, list number of days exceeded for each sample:   |    |    |     |

ADDITIONAL COMMENTS:

\_\_\_\_\_  
QA REVIEW

\_\_\_\_\_  
Date