

TitriSoft Report

Reviewed By: Iwona
On: 11/20/2024 2:07:27 PM
Inst Id: Titroline Alpha
Titrator

Ident	Amount	DateTime	Alkalinity	Final pH (sample)	Initial pH (sample)	slope	offset
LB133515BLW	100	19/11/2024 10:15	0.00	3.89	3.92	102.8	6.98
LB133515BSW	100	19/11/2024 10:20	43.56	4.34	8.62	102.8	6.98
P4846-01	50	19/11/2024 10:28	192.80	4.41	7.27	102.8	6.98
P4846-01DUP	50	19/11/2024 10:36	192.76	4.42	7.26	102.8	6.98
P4846-02	50	19/11/2024 10:44	241.64	4.45	7.40	102.8	6.98
P4846-03	50	19/11/2024 10:52	241.24	4.50	7.37	102.8	6.98
P4846-04	50	19/11/2024 11:00	230.48	4.38	7.40	102.8	6.98
P4846-05	50	19/11/2024 11:08	224.48	4.46	7.37	102.8	6.98
P4846-08	50	19/11/2024 11:15	197.26	4.40	7.45	102.8	6.98
P4846-10	50	19/11/2024 11:20	2.00	4.38	5.50	102.8	6.98
P4846-11	50	19/11/2024 11:26	190.82	4.42	7.44	102.8	6.98
P4846-12	50	19/11/2024 11:34	245.64	4.48	7.36	102.8	6.98
P4846-13	50	19/11/2024 11:40	153.68	4.41	7.53	102.8	6.98
P4846-14	50	19/11/2024 11:46	284.62	4.42	7.16	102.8	6.98
P4846-15	50	19/11/2024 11:54	287.54	4.43	7.11	102.8	6.98
P4846-16	50	19/11/2024 12:00	172.48	4.42	7.24	102.8	6.98
P4846-17	50	19/11/2024 12:08	237.62	4.46	7.19	102.8	6.98
P4873-01	50	19/11/2024 12:12	2.00	4.43	5.10	102.8	6.98
P4873-02	50	19/11/2024 12:18	28.42	4.34	7.17	102.8	6.98

11/19/2024

11/19/24
NF

WORKLIST(Hardcopy Internal Chain)

LB 133515

WorkList Name : ALKALANITY-111924

WorkList ID : 185575

Department : Wet-Chemistry

Date : 11-19-2024 09:47:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4846-01	FSND-MW-EVAL-03D-20241111	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/13/2024	SM2320 B
P4846-02	FSND-MW-EVAL-03S-20241111	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/13/2024	SM2320 B
P4846-03	FSND-FD-3-20241113	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/13/2024	SM2320 B
P4846-04	FSND-MW-EVAL-04S-20241111	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/13/2024	SM2320 B
P4846-05	FSND-MW-EVAL-04D-20241111	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/13/2024	SM2320 B
P4846-08	FSND-MW-26-20241113	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/13/2024	SM2320 B
P4846-10	FSND-RB-3-20241113	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/13/2024	SM2320 B
P4846-11	FSND-MW-24-20241113	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/13/2024	SM2320 B
P4846-12	FSND-MW-15-20241113	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/13/2024	SM2320 B
P4846-13	FSND-MW-EVAL-02D-20241111	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/13/2024	SM2320 B
P4846-14	FSND-MW-EVAL-02S-20241111	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/13/2024	SM2320 B
P4846-15	FSND-FD-4-20241113	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/13/2024	SM2320 B
P4846-16	FSND-MW-EVAL-01D-20241111	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/13/2024	SM2320 B
P4846-17	FSND-MW-EVAL-01S-20211113	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/13/2024	SM2320 B
P4873-01	FSND-RB-4-20241114	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/13/2024	SM2320 B
P4873-02	FSND-MW-30-20241114	Water	Alkalinity	Cool 4 deg C	EAEN05	M11	11/14/2024	SM2320 B
					EAEN05	M11	11/14/2024	SM2320 B

Date/Time 11/19/2024 10:00
 Raw Sample Received by: NF(wc)
 Raw Sample Relinquished by: NF(wc)

Date/Time 11/19/2024 13:00
 Raw Sample Received by: NF(wc)
 Raw Sample Relinquished by: NF(wc)